



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85367

Temp.: 14.6 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	01/17/2017	11:00
2	CAL2	1	Water			7.01	01/17/2017	11:01
3	CAL3	1	Water			10.03	01/17/2017	11:02
4	ICV	1	Water			7.02	01/17/2017	11:03
5	CCV1	1	Water			2.01	01/17/2017	11:05
6	I1248-01	1	Water			3.99	01/17/2017	11:06
7	I1248-01DUP	1	Water			4.00	01/17/2017	11:07
8	I1248-02	1	Water			4.19	01/17/2017	11:08
9	I1248-03	1	Water			5.64	01/17/2017	11:09
10	CCV2	1	Water			12.02	01/17/2017	11:10

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 11248

WorkList ID : 94484

Date : 1/17/2017 10:27:44 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1248-01	pH	Cool 4 deg C	STER09	D41	CELL-A-POOLWATER	01/16/2017	SM 4500-PH B
	Water	I1248-02	pH	Cool 4 deg C	STER09	D41	CELL-B-C-POOLWATER	01/16/2017	SM 4500-PH B
	Water	I1248-03	pH	Cool 4 deg C	STER09	D41	CELL-D-POOLWATER	01/16/2017	SM 4500-PH B

Date/Time 01-17-17 10:27 AM
 Received by: JP
 Relinquished by: JP

Date/Time 01-17-17 11:40 AM
 Received by: JP
 Relinquished by: JP