



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: JIGNESHDate: 01/18/2017Run Number: LB85374DATE & TIME IN: 01/17/2017 14:30OVEN TEMP: 104 °CDATE & TIME OUT: 01/18/2017 18:35OVEN TEMP: 105 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
1	LB85374BL	LB85374BL	1.5237	1.5237	100	1.5237	1.5237	1.5237	0	0
2	LB85374BS	LB85374BS	1.4602	1.4602	100	1.5146	1.5146	1.5146	0.0544	544
3	I1244-01	001M-WILLETS-PTBLVD (JAN)	1.4127	1.4127	100	1.4458	1.4458	1.4458	0.0331	331
4	I1244-02	002M-35THAVE (JAN)	1.4081	1.4081	100	1.4644	1.4644	1.4644	0.0563	563
5	I1244-02DUP	002M-35THAVE (JAN) DUP	1.4175	1.4175	100	1.4742	1.4742	1.4742	0.0567	567

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

1885374

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS I1244

WorkList ID : 94497

Date : 1/17/2017 1:50:37 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1244-01	TSS	Cool 4 deg C	TULL01	D41	001M-WILLETS-PTBLVD(JAN	01/16/2017	SM2540 D
	Water	I1244-02	TSS	Cool 4 deg C	TULL01	D41	002M-35THAVE(JAN)	01/16/2017	SM2540 D

Date/Time 01.17.17 11:50 AM
Received by: JH
Relinquished by: CP

Date/Time 01.17.17 09:00 AM
Received by: CP
Relinquished by: JH