


Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/29/17
Analyst : AP
Data File : LB85382.csv

Approved By : 
Approved Date :
Worksheet # : 45018

M300-0 - Inorganic Arsenic

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Nitrite												
STD1	STD1				W	1/17/17						
Nitrite		PASS	0.000		0.000						mg/L	
STD2	STD2				W	1/17/17						
Nitrite		PASS	0.331		0.331						mg/L	
STD3	STD3				W	1/17/17						
Nitrite		PASS	0.682		0.682						mg/L	
STD4	STD4				W	1/17/17						
Nitrite		PASS	1.224		1.22						mg/L	
STD5	STD5				W	1/17/17						
Nitrite		PASS	2.877		2.88						mg/L	
STD6	STD6				W	1/17/17						
Nitrite		PASS	5.868		5.87						mg/L	
STD7	STD7				W	1/17/17						
Nitrite		PASS	7.415		7.42						mg/L	
STD8	STD8				W	1/17/17						
Nitrite		PASS	15.363		15.4						mg/L	
CCV1	CCV1				W	1/17/17						
Nitrite		PASS	7.328		7.33	96.0	90	110			mg/L	
CCB1	CCB1				W	1/17/17						
Nitrite		PASS	0.000		0.000			+/-0.0220			mg/L	
ICV1	ICV1				W	1/18/17						
Nitrite		PASS	7.389		7.39	99.0	90	110			mg/L	
ICB1	ICB1				W	1/18/17						
Nitrite		PASS	0.000		0.000			+/-0.0220			mg/L	
CCV2	CCV2				W	1/20/17						
Nitrite		PASS	7.151		7.15	94.0	90	110			mg/L	
CCB2	CCB2				W	1/20/17						
Nitrite		PASS	0.000		0.000			+/-0.0220			mg/L	
LB85382BLW	LB85382BLW				W	1/20/17						
Nitrite		PASS	0.000		0.000			+/-0.0220			mg/L	
LB85382BSW	LB85382BSW				W	1/20/17						
Nitrite		PASS	7.530		7.53	100.0	90	110			mg/L	
I1002-08	MDL-WATER-1		1		W	1/20/17						
Nitrite		PASS	0.085		0.085						mg/L	
I1002-09	MDL-WATER-2		1		W	1/20/17						
Nitrite		PASS	0.086		0.086						mg/L	
I1002-10	MDL-WATER-3		1		W	1/20/17						
Nitrite		PASS	0.086		0.086						mg/L	
I1002-11	MDL-WATER-4		1		W	1/20/17						
Nitrite		PASS	0.086		0.086						mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/29/17

Analyst : AP

Data File : LB85382.csv

Approved By : 

Approved Date : 1/29/17

Worksheet # : 120/10

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Nitrite											
I1002-12	MDL-WATER-5		1	W	1/20/17						
Nitrite	PASS	0.086		0.086						mg/L	
I1002-13	MDL-WATER-6		1	W	1/20/17						
Nitrite	PASS	0.085		0.085						mg/L	
I1002-14	MDL-WATER-7		1	W	1/20/17						
Nitrite	PASS	0.086		0.086						mg/L	
I1002-14	MDL-WATER-7		1	W	1/20/17						
Nitrite	PASS	0.087		0.087						mg/L	
CCV3	CCV3			W	1/20/17						
Nitrite	PASS	7.312		7.31	96.0	90	110			mg/L	
CCB3	CCB3			W	1/20/17						
Nitrite	PASS	0.000		0.000			+/-0.0220			mg/L	