


7470A

LB 85383
INSTRUMENT ID: CV1

Linear

 μ Abs.:

38000



A= 0.0000e+000

B= 2.5936e-004

C= -1.4291e-003

Rho= 0.9995274

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.000	-0.000	5	0.000	5				
0.2	0.200	0.175	-0.025	682	0.0 %	682				
2.5	2.500	2.512	0.012	9690	0.0 %	9690				
5.0	5.000	4.934	-0.066	19028	0.0 %	19028				
7.5	7.500	7.725	0.225	29792	0.0 %	29792				
10.0	10.000	9.854	-0.146	38000	0.0 %	38000				

LB85383 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Jan 2017 17:05:03

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	5	-	0.00007470A	PPB	17 Jan 2017 17:09:34	S	Std
0.2 - 1	50.2	682	-	0.20007470A	PPB	17 Jan 2017 17:11:42	S	Std
2.5 - 1	52.5	9690	-	2.50007470A	PPB	17 Jan 2017 17:13:51	S	Std
5.0 - 1	55	19028	-	5.00007470A	PPB	17 Jan 2017 17:15:59	S	Std
7.5 - 1	57.5	29792	-	7.50007470A	PPB	17 Jan 2017 17:18:07	S	Std
10.0 - 1	50	38000	-	10.00007470A	PPB	17 Jan 2017 17:20:15	S	Std
ICV99 - 1	ICV99	14747	3.8233	-7470A	PPB	17 Jan 2017 17:22:49	U	SMPL
ICB99 - 1	ICB99	-29	-0.0090	-7470A	PPB	17 Jan 2017 17:24:55	U	SMPL
CCV69 - 1	CCV69	18854	4.8885	-7470A	PPB	17 Jan 2017 17:27:01	U	SMPL
CCB69 - 1	CCB69	-27	-0.0084	-7470A	PPB	17 Jan 2017 17:29:08	U	SMPL
CRA - 1	CRA	644	0.1656	-7470A	PPB	17 Jan 2017 17:31:15	U	SMPL
HighStd - 1	HighStd	38278	9.9262	-7470A	PPB	17 Jan 2017 17:33:21	U	SMPL
ChkStd - 1	ChkStd	26225	6.8002	-7470A	PPB	17 Jan 2017 17:35:28	U	SMPL
PB96030BL - 1	PBW	0	-0.0014	-7470A	PPB	17 Jan 2017 17:37:35	U	SMPL
PB96030TB - 1	PB95984TB	-9	-0.0038	-7470A	PPB	17 Jan 2017 17:41:50	U	SMPL
I1182-05 - 1	TOD-WCS-011217	0	-0.0014	-7470A	PPB	17 Jan 2017 17:43:59	U	SMPL
I1182-05DUP - 1	TOD-WCS-011217DUP	8	0.0006	-7470A	PPB	17 Jan 2017 17:46:07	U	SMPL
I1182-05MS - 1	TOD-WCS-011217MS	14947	3.8752	-7470A	PPB	17 Jan 2017 17:50:40	U	SMPL
I1182-05MSD - 1	TOD-WCS-011217MSD	14876	3.8567	-7470A	PPB	17 Jan 2017 17:52:46	U	SMPL
CCV70 - 1	CCV70	18646	4.8345	-7470A	PPB	17 Jan 2017 17:54:53	U	SMPL
CCB70 - 1	CCB70	-22	-0.0071	-7470A	PPB	17 Jan 2017 17:56:59	U	SMPL
I1219-01 - 1	TOD-WCS-011317(11-47)	14	0.0022	-7470A	PPB	17 Jan 2017 17:59:06	U	SMPL
PB96035BL - 1	PBW	12	0.0017	-7470A	PPB	17 Jan 2017 18:01:13	U	SMPL
PB96035BS - 1	LCSW	15330	3.9745	-7470A	PPB	17 Jan 2017 18:03:20	U	SMPL
I1182-06 - 1	FB011217-TOD	-20	-0.0066	-7470A	PPB	17 Jan 2017 18:05:27	U	SMPL
I1219-02 - 1	WC011217-LIQUID-TODT-HILL	336	0.0857	-7470A	PPB	17 Jan 2017 18:07:35	U	SMPL
I1219-02DUP - 1	WC011217-LIQUID-TODT-HILLDUP	334	0.0852	-7470A	PPB	17 Jan 2017 18:09:42	U	SMPL
I1219-02MS - 1	WC011217-LIQUID-TODT-HILLMS	15334	3.9755	-7470A	PPB	17 Jan 2017 18:11:51	U	SMPL
I1219-02MSD - 1	WC011217-LIQUID-TODT-HILLMSD	15379	3.9872	-7470A	PPB	17 Jan 2017 18:13:59	U	SMPL
CCV71 - 1	CCV71	18965	4.9172	-7470A	PPB	17 Jan 2017 18:16:08	U	SMPL
CCB71 - 1	CCB71	-23	-0.0074	-7470A	PPB	17 Jan 2017 18:18:15	U	SMPL
I1182-05LX5 - 1		9	0.0009	-7470A	PPB	17 Jan 2017 18:20:22	U	SMPL
I1182-05A - 1		14651	3.7984	-7470A	PPB	17 Jan 2017 18:22:28	U	SMPL
I1219-02LX5 - 1		51	0.0118	-7470A	PPB	17 Jan 2017 18:24:35	U	SMPL
I1219-02A - 1		14698	3.8106	-7470A	PPB	17 Jan 2017 18:26:43	U	SMPL
PB96030BS - 1	LCSW	14892	3.8609	-7470A	PPB	17 Jan 2017 18:28:49	U	SMPL
CCV72 - 1	CCV72	18974	4.9196	-7470A	PPB	17 Jan 2017 18:30:56	U	SMPL
CCB72 - 1	CCB72	-27	-0.0084	-7470A	PPB	17 Jan 2017 18:33:04	U	SMPL