

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17
 Analyst : AP
 Data File : LB85384.csv

Approved By : 
 Approved Date : 1/17/17
 Worksheet # :

M300.0 - Inorganic Ammonia - 16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Bromide												
STD1	STD1				W	1/16/17						
Bromide		PASS	0.000		0.000						mg/L	
STD2	STD2				W	1/16/17						
Bromide		PASS	0.913		0.913						mg/L	
STD3	STD3				W	1/16/17						
Bromide		PASS	2.270		2.27						mg/L	
STD4	STD4				W	1/16/17						
Bromide		PASS	4.082		4.08						mg/L	
STD5	STD5				W	1/16/17						
Bromide		PASS	9.756		9.76						mg/L	
STD6	STD6				W	1/16/17						
Bromide		PASS	19.527		19.5						mg/L	
STD7	STD7				W	1/16/17						
Bromide		PASS	24.510		24.5						mg/L	
STD8	STD8				W	1/16/17						
Bromide		PASS	50.442		50.4						mg/L	
ICV1	ICV1				W	1/16/17						
Bromide		PASS	25.395		25.40	102.0	90	110			mg/L	
ICB1	ICB1				W	1/16/17						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	
CCV1	CCV1				W	1/17/17						
Bromide		PASS	24.991		25.00	100.0	90	110			mg/L	
CCB1	CCB1				W	1/17/17						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	
LB85384BLW	LB85384BLW				W	1/17/17						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	
LB85384BSW	LB85384BSW				W	1/17/17						
Bromide		PASS	25.305		25.30	101.0	90	110			mg/L	
I1248-01	CELL-A-POOLWATER		1		W	1/17/17						
Bromide		PASS	0.000		0.000						mg/L	
I1248-02	CELL-B-C-POOLWATER		1		W	1/17/17						
Bromide		PASS	0.000		0.000						mg/L	
I1248-03	CELL-D-POOLWATER		1		W	1/17/17						
Bromide		PASS	0.000		0.000						mg/L	
I1248-03D	CELL-D-POOLWATERD		1		W	1/17/17						
Bromide		PASS	0.000		0.00				0	20	mg/L	
I1248-03S	CELL-D-POOLWATERS		1		W	1/17/17						
Bromide		PASS	24.972		25.00	100.0	80	120			mg/L	
CCV2	CCV2				W	1/17/17						
Bromide		PASS	25.486		25.50	102.0	90	110			mg/L	


Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17

Analyst : AP

Data File : LB85384.csv

Approved By : 

Approved Date : 1/17/17

Worksheet # : 1

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis					Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units		Line 2
Bromide												
CCB2	CCB2			W	1/17/17							
Bromide		PASS		0.000	0.000		+/-0.0660			mg/L		

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17
 Analyst : AP
 Data File : LB85384.csv

Approved By : 
 Approved Date :
 Worksheet # : 11912

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method	LCL	UCL	Line 2
Chloride										
STD1	STD1			W	1/16/17					
Chloride		PASS		0.000	0.000					mg/L
STD2	STD2			W	1/16/17					
Chloride		PASS		0.357	0.357					mg/L
STD3	STD3			W	1/16/17					
Chloride		PASS		0.765	0.765					mg/L
STD4	STD4			W	1/16/17					
Chloride		PASS		1.259	1.26					mg/L
STD5	STD5			W	1/16/17					
Chloride		PASS		2.867	2.87					mg/L
STD6	STD6			W	1/16/17					
Chloride		PASS		5.722	5.72					mg/L
STD7	STD7			W	1/16/17					
Chloride		PASS		7.248	7.25					mg/L
STD8	STD8			W	1/16/17					
Chloride		PASS		15.231	15.2					mg/L
ICV1	ICV1			W	1/16/17					
Chloride		PASS		7.295	7.30	97.0		90	110	mg/L
ICB1	ICB1			W	1/16/17					
Chloride		PASS		0.000	0.000				+/-0.0750	mg/L
CCV1	CCV1			W	1/17/17					
Chloride		PASS		7.250	7.25	97.0		90	110	mg/L
CCB1	CCB1			W	1/17/17					
Chloride		PASS		0.000	0.000				+/-0.0750	mg/L
LB85384BLW	LB85384BLW			W	1/17/17					
Chloride		PASS		0.000	0.000				+/-0.0750	mg/L
LB85384BSW	LB85384BSW			W	1/17/17					
Chloride		PASS		7.424	7.42	99.0		90	110	mg/L
I1248-01	CELL-A-POOLWATER		1	W	1/17/17					
Chloride		PASS		0.000	0.000					mg/L
I1248-02	CELL-B-C-POOLWATER		1	W	1/17/17					
Chloride		PASS		0.000	0.000					mg/L
I1248-03	CELL-D-POOLWATER		1	W	1/17/17					
Chloride		PASS		0.567	0.567					mg/L
I1248-03D	CELL-D-POOLWATERD		1	W	1/17/17					
Chloride		PASS		0.554	0.55				3	20
I1248-03S	CELL-D-POOLWATERS		1	W	1/17/17					
Chloride		PASS		7.421	7.42	91.0		80	120	mg/L
CCV2	CCV2			W	1/17/17					
Chloride		PASS		7.555	7.56	101.0		90	110	mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17
Analyst : AP
Data File : LB85384.csv


Approved By : 
Approved Date : 1/17/17
Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1		
Parameter		PPB		Final Conc	%Rec	Method	Method	LCL	UCL	RPD	Max RPD	Units	Line 2
Chloride													
CCB2	CCB2			W	1/17/17								
Chloride		PASS		0.000	0.000				+/-0.0750			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17
 Analyst : AP
 Data File : LB85384.csv

Approved By : 
 Approved Date : 
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1	
Parameter			PPB		Final Conc	%Rec	Method	LCL	UCL	RPD	Max RPD	Units	Line 2
Fluoride													
STD1	STD1				W	1/16/17							
Fluoride		PASS	0.000		0.000							mg/L	
STD2	STD2				W	1/16/17							
Fluoride		PASS	0.255		0.255							mg/L	
STD3	STD3				W	1/16/17							
Fluoride		PASS	0.470		0.470							mg/L	
STD4	STD4				W	1/16/17							
Fluoride		PASS	0.845		0.845							mg/L	
STD5	STD5				W	1/16/17							
Fluoride		PASS	1.854		1.85							mg/L	
STD6	STD6				W	1/16/17							
Fluoride		PASS	3.888		3.89							mg/L	
STD7	STD7				W	1/16/17							
Fluoride		PASS	4.862		4.86							mg/L	
STD8	STD8				W	1/16/17							
Fluoride		PASS	10.125		10.1							mg/L	
ICV1	ICV1				W	1/16/17							
Fluoride		PASS	4.693		4.69	94.0	90	110				mg/L	
ICB1	ICB1				W	1/16/17							
Fluoride		PASS	0.000		0.000			+/-0.0430				mg/L	
CCV1	CCV1				W	1/17/17							
Fluoride		PASS	4.982		4.98	100.0	90	110				mg/L	
CCB1	CCB1				W	1/17/17							
Fluoride		PASS	0.000		0.000			+/-0.0430				mg/L	
LB85384BLW	LB85384BLW				W	1/17/17							
Fluoride		PASS	0.000		0.000			+/-0.0430				mg/L	
LB85384BSW	LB85384BSW				W	1/17/17							
Fluoride		PASS	4.928		4.93	99.0	90	110				mg/L	
I1248-01	CELL-A-POOLWATER		1		W	1/17/17							
Fluoride		PASS	0.000		0.000							mg/L	
I1248-02	CELL-B-C-POOLWATER		1		W	1/17/17							
Fluoride		PASS	0.000		0.000							mg/L	
I1248-03	CELL-D-POOLWATER		1		W	1/17/17							
Fluoride		PASS	0.000		0.000							mg/L	
I1248-03D	CELL-D-POOLWATERD		1		W	1/17/17							
Fluoride		PASS	0.000		0.00					0	20	mg/L	
I1248-03S	CELL-D-POOLWATERS		1		W	1/17/17							
Fluoride		PASS	4.804		4.80	96.0	80	120				mg/L	
CCV2	CCV2				W	1/17/17							
Fluoride		PASS	5.052		5.05	101.0	90	110				mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17
Analyst : AP
Data File : LB85384.csv

Approved By : 
Approved Date : 1/17/17
Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method				Line 2
Fluoride											
CCB2	CCB2			W	1/17/17						
Fluoride		PASS	0.000	0.000			+/-0.0430			mg/L	