


7471A

LAB 85388
INSTRUMENT ID: CV1

Linear

 μ Abs.:

46070



A= 0.0000e+000

B= 2.1745e-004

C= 1.4094e-002

Rho= 0.9999720

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	-17	0.000	-17				
0.2	0.200	0.219	0.019	941	0.0 %	941				
2.5	2.500	2.502	0.002	11439	0.0 %	11439				
5.0	5.000	4.945	-0.055	22676	0.0 %	22676				
7.5	7.500	7.492	-0.008	34390	0.0 %	34390				
10.0	10.000	10.032	0.032	46070	0.0 %	46070				

LB85388 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 18 Jan 2017 11:25:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-17	-	0.0000	7471A	PPB	18 Jan 2017 11:27:12	S	Std
0.2 - 1	50.2	941	-	0.2000	7471A	PPB	18 Jan 2017 11:29:30	S	Std
2.5 - 1	52.5	11439	-	2.5000	7471A	PPB	18 Jan 2017 11:31:48	S	Std
5.0 - 1	55	22676	-	5.0000	7471A	PPB	18 Jan 2017 11:34:05	S	Std
7.5 - 1	57.5	34390	-	7.5000	7471A	PPB	18 Jan 2017 11:36:23	S	Std
10.0 - 1	510	46070	-	10.0000	7471A	PPB	18 Jan 2017 11:41:23	S	Std
ICV01 - 1	ICV01	17526	3.8251	-	7471A	PPB	18 Jan 2017 11:44:01	U	SMPL
ICB01 - 1	ICB01	-37	0.0060	-	7471A	PPB	18 Jan 2017 11:46:17	U	SMPL
CCV73 - 1	CCV73	22210	4.8437	-	7471A	PPB	18 Jan 2017 11:48:32	U	SMPL
CCB73 - 1	CCB73	-43	0.0047	-	7471A	PPB	18 Jan 2017 11:50:48	U	SMPL
CRA - 1	CRA	929	0.2161	-	7471A	PPB	18 Jan 2017 11:53:04	U	SMPL
HighStd - 1	HighStd	47219	10.2819	-	7471A	PPB	18 Jan 2017 11:55:19	U	SMPL
ChkStd - 1	ChkStd	31789	6.9267	-	7471A	PPB	18 Jan 2017 11:57:36	U	SMPL
PB96073BL - 1	PBS	-46	0.0041	-	7471A	PPB	18 Jan 2017 11:59:52	U	SMPL
PB96073BS - 1	LCSS	17532	3.8265	-	7471A	PPB	18 Jan 2017 12:04:43	U	SMPL
I1182-01 - 1	TOD-BC-0.2	11909	2.6037	-	7471A	PPB	18 Jan 2017 12:06:59	U	SMPL
I1182-03 - 1	TOD-SB-23-25	99	0.0356	-	7471A	PPB	18 Jan 2017 12:09:17	U	SMPL
I1182-04 - 1	TOD-SB-78-80	82	0.0319	-	7471A	PPB	18 Jan 2017 12:11:34	U	SMPL
I1243-01 - 1	PHOENIX-BPM	171	0.0513	-	7471A	PPB	18 Jan 2017 12:13:52	U	SMPL
I1243-01DUP - 1	PHOENIX-BPMDUP	217	0.0613	-	7471A	PPB	18 Jan 2017 12:16:08	U	SMPL
CCV74 - 1	CCV74	22932	5.0007	-	7471A	PPB	18 Jan 2017 12:22:56	U	SMPL
CCB74 - 1	CCB74	-47	0.0039	-	7471A	PPB	18 Jan 2017 12:25:11	U	SMPL
I1243-01MS - 1	PHOENIX-BPMMS	17862	3.8982	-	7471A	PPB	18 Jan 2017 12:27:27	U	SMPL
I1243-01MSD - 1	PHOENIX-BPMMSD	17841	3.8936	-	7471A	PPB	18 Jan 2017 12:29:43	U	SMPL
I1246-01 - 1	ROW-SP-0-2-COMP	2624	0.5847	-	7471A	PPB	18 Jan 2017 12:32:00	U	SMPL
I1243-01LX5 - 1		19	0.0182	-	7471A	PPB	18 Jan 2017 12:34:16	U	SMPL
I1243-01A - 1		17789	3.8823	-	7471A	PPB	18 Jan 2017 12:36:33	U	SMPL
CCV75 - 1	CCV75	22718	4.9542	-	7471A	PPB	18 Jan 2017 12:38:50	U	SMPL
CCB75 - 1	CCB75	-36	0.0063	-	7471A	PPB	18 Jan 2017 12:41:07	U	SMPL