



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB85391

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	01/18/2017	12:55
2	I1219-03		1	>100.0	>212.0	01/18/2017	13:20
3	I1219-03DUP		1	>100.0	>212.0	01/18/2017	13:40

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP011817

WorkList ID : 94563

Date : 1/18/2017 12:17:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/26/2017	Water	I1219-03	Flash Point	Cool 4 deg C	PARS15	D32	WC011217-LIQUID-TODT-HILL01/12/2017		1010A

Date/Time 1/19/17 12:30P

Received by: AS

Relinquished by: AS

Date/Time 1/19/17 2:30P

Received by: AS

Relinquished by: AS