

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/19/17

Analyst : AP

Data File : LB85393.csv

Approved By : 

Approved Date : 1/20/17

Worksheet # :

M300.0 - Inorganic Ammonia - 16

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
Sulfate										
STD1	STD1			W	1/16/17					
Sulfate		PASS	0.000	0.000						mg/L
STD2	STD2			W	1/16/17					
Sulfate		PASS	2.257	2.26						mg/L
STD3	STD3			W	1/16/17					
Sulfate		PASS	4.027	4.03						mg/L
STD4	STD4			W	1/16/17					
Sulfate		PASS	6.435	6.44						mg/L
STD5	STD5			W	1/16/17					
Sulfate		PASS	14.161	14.2						mg/L
STD6	STD6			W	1/16/17					
Sulfate		PASS	28.134	28.1						mg/L
STD7	STD7			W	1/16/17					
Sulfate		PASS	35.683	35.7						mg/L
STD8	STD8			W	1/16/17					
Sulfate		PASS	76.553	76.6						mg/L
ICV1	ICV1			W	1/16/17					
Sulfate		PASS	36.442	36.40	97.0	90	110			mg/L
ICB1	ICB1			W	1/16/17					
Sulfate		PASS	0.000	0.000			+/-0.1320			mg/L
CCV1	CCV1			W	1/18/17					
Sulfate		PASS	35.673	35.70	95.0	90	110			mg/L
CCB1	CCB1			W	1/18/17					
Sulfate		PASS	0.000	0.000			+/-0.1320			mg/L
LB85393BLW	LB85393BLW			W	1/18/17					
Sulfate		PASS	0.000	0.000			+/-0.1320			mg/L
LB85393BSW	LB85393BSW			W	1/18/17					
Sulfate		PASS	37.087	37.10	99.0	90	110			mg/L
I1275-01	MW1		1	W	1/18/17					
Sulfate		OR	107.000	107						mg/L
I1275-02	MW3		1	W	1/18/17					
Sulfate		PASS	3.774	3.77						mg/L
I1275-03	MW4		1	W	1/18/17					
Sulfate		PASS	14.677	14.7						mg/L
I1275-04	MW5		1	W	1/18/17					
Sulfate		PASS	17.383	17.4						mg/L
I1275-05	MW10		1	W	1/18/17					
Sulfate		PASS	2.822	2.82						mg/L
I1275-06	MW11		1	W	1/18/17					
Sulfate		PASS	2.758	2.76						mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/19/17

Analyst : AP

Data File : LB85393.csv

Approved By : 

Approved Date :

Worksheet # : yw/b

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
Sulfate											
I1275-06D	MW11D			1	W	1/18/17					
Sulfate		PASS	2.763		2.76				0	20	mg/L
I1275-06S	MW11S			1	W	1/18/17					
Sulfate		PASS	37.507		37.50	93.0	80	120			mg/L
CCV2	CCV2				W	1/18/17					
Sulfate		PASS	36.066		36.10	96.0	90	110			mg/L
CCB2	CCB2				W	1/18/17					
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L
I1275-01DL	MW1DL			2	W	1/18/17					
Sulfate		PASS	47.194		94.4						mg/L
CCV3	CCV3				W	1/18/17					
Sulfate		PASS	36.744		36.70	98.0	90	110			mg/L
CCB3	CCB3				W	1/18/17					
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L