

7470A


L385400

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39383



A= 0.0000e+000

B= 2.5554e-004

C= -1.2139e-002

Rho= 0.9996942

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	29	0.000	29				
0.2	0.200	0.193	-0.007	803	0.0 %	803				
2.5	2.500	2.482	-0.018	9762	0.0 %	9762				
5.0	5.000	5.141	0.141	20167	0.0 %	20167				
7.5	7.500	7.336	-0.164	28755	0.0 %	28755				
10.0	10.000	10.052	0.052	39383	0.0 %	39383				

LB85400 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Jan 2017 16:01:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	29	-	0.0000	7470A	PPB	18 Jan 2017 16:03:29	S	Std
0.2 - 1	S0-2	803	-	0.2000	7470A	PPB	18 Jan 2017 16:05:37	S	Std
2.5 - 1	S2-5	9762	-	2.5000	7470A	PPB	18 Jan 2017 16:07:45	S	Std
5.0 - 1	S5	20167	-	5.0000	7470A	PPB	18 Jan 2017 16:09:53	S	Std
7.5 - 1	S7.5	28755	-	7.5000	7470A	PPB	18 Jan 2017 16:12:01	S	Std
10.0 - 1	S10	39383	-	10.0000	7470A	PPB	18 Jan 2017 16:14:08	S	Std
ICV02 - 1	ICV02	14264	3.6329	-	7470A	PPB	18 Jan 2017 16:16:49	U	SMPL
ICB02 - 1	ICB02	13	-0.0088	-	7470A	PPB	18 Jan 2017 16:18:55	U	SMPL
CCV76 - 1	CCV76	19823	5.0535	-	7470A	PPB	18 Jan 2017 16:21:01	U	SMPL
CCB76 - 1	CCB76	5	-0.0109	-	7470A	PPB	18 Jan 2017 16:23:08	U	SMPL
CRA - 1	CRA	788	0.1892	-	7470A	PPB	18 Jan 2017 16:25:14	U	SMPL
HighStd - 1	HighStd	38863	9.9190	-	7470A	PPB	18 Jan 2017 16:27:21	U	SMPL
ChkStd - 1	ChkStd	26753	6.8244	-	7470A	PPB	18 Jan 2017 16:29:28	U	SMPL
PB96076BL - 1	PBW	-10	-0.0147	-	7470A	PPB	18 Jan 2017 16:31:35	U	SMPL
PB96076BS - 1	LCSW	14796	3.7689	-	7470A	PPB	18 Jan 2017 16:33:43	U	SMPL
PB96076TB - 1	PB96036TB	-2	-0.0126	-	7470A	PPB	18 Jan 2017 16:35:50	U	SMPL
I1246-03 - 1	ROW-SP-0-2-COMP	31	-0.0042	-	7470A	PPB	18 Jan 2017 16:37:58	U	SMPL
I1246-03DUP - 1	ROW-SP-0-2-COMPDUP	33	-0.0037	-	7470A	PPB	18 Jan 2017 16:40:06	U	SMPL
I1246-03MS - 1	ROW-SP-0-2-COMPMS	14857	3.7844	-	7470A	PPB	18 Jan 2017 16:42:15	U	SMPL
I1246-03MSD - 1	ROW-SP-0-2-COMPMSD	14580	3.7137	-	7470A	PPB	18 Jan 2017 16:44:22	U	SMPL
CCV77 - 1	CCV77	19633	5.0049	-	7470A	PPB	18 Jan 2017 16:46:29	U	SMPL
CCB77 - 1	CCB77	-4	-0.0132	-	7470A	PPB	18 Jan 2017 16:48:35	U	SMPL
I1246-03LX5 - 1		23	-0.0063	-	7470A	PPB	18 Jan 2017 16:50:42	U	SMPL
I1246-03A - 1		14786	3.7663	-	7470A	PPB	18 Jan 2017 16:52:49	U	SMPL
CCV78 - 1	CCV78	19349	4.9323	-	7470A	PPB	18 Jan 2017 16:54:56	U	SMPL
CCB78 - 1	CCB78	2	-0.0116	-	7470A	PPB	18 Jan 2017 16:57:03	U	SMPL