

ISM02.3


LB85404

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45358



A= 0.0000e+000

B= 2.1882e-004

C= 1.7305e-002

Rho= 0.9993507

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.022	0.022	22	0.000	22					15
0.20	0.200	0.230	0.030	970	0.0 %	970					-8
2.50	2.500	2.295	-0.205	10410	0.0 %	10410					5
5.00	5.000	5.242	0.242	23876	0.0 %	23876					0
7.50	7.500	7.469	-0.031	34053	0.0 %	34053					-1
10.00	10.000	9.943	-0.057	45358	0.0 %	45358					

NB
1/18/17

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	22	-		0 ISM02.3	PPB	1/18/2017 17:13	Std
	0.2 S0.2	970	-		0.2 ISM02.3	PPB	1/18/2017 17:15	Std
	2.5 S2.5	10410	-		2.5 ISM02.3	PPB	1/18/2017 17:17	Std
	5 S5	23876	-		5 ISM02.3	PPB	1/18/2017 17:19	Std
	7.5 S7.5	34053	-		7.5 ISM02.3	PPB	1/18/2017 17:26	Std
	10 S10	45358	-		10 ISM02.3	PPB	1/18/2017 17:28	Std
ICV	ICV	18463	4.0574	-	ISM02.3	PPB	1/18/2017 17:32	SMPL
ICB	ICB	4	0.0182	-	ISM02.3	PPB	1/18/2017 17:34	SMPL
CCV079	CCV079	23828	5.2313	-	ISM02.3	PPB	1/18/2017 17:36	SMPL
CCB079	CCB079	-2	0.0169	-	ISM02.3	PPB	1/18/2017 17:38	SMPL
PB96074BL	PBS004	29	0.0237	-	ISM02.3	PPB	1/18/2017 17:40	SMPL
I1133-01	ME5BL5	6180	1.3696	-	ISM02.3	PPB	1/18/2017 17:42	SMPL
I1133-02	ME5BL6	2636	0.5941	-	ISM02.3	PPB	1/18/2017 17:44	SMPL
I1133-03	ME5BL7	3531	0.79	-	ISM02.3	PPB	1/18/2017 17:46	SMPL
I1133-04	ME5BL8	2173	0.4928	-	ISM02.3	PPB	1/18/2017 17:49	SMPL
I1133-05	ME5BL9	2396	0.5416	-	ISM02.3	PPB	1/18/2017 17:51	SMPL
I1133-06	ME5BM0	2799	0.6298	-	ISM02.3	PPB	1/18/2017 17:53	SMPL
I1133-07	ME5BM1	4970	1.1048	-	ISM02.3	PPB	1/18/2017 17:55	SMPL
I1133-08	ME5BM2	4098	0.914	-	ISM02.3	PPB	1/18/2017 17:57	SMPL
I1133-09	ME5BM3	4542	1.0112	-	ISM02.3	PPB	1/18/2017 17:59	SMPL
I1133-10	ME5BM4	11096	2.4453	-	ISM02.3	PPB	1/18/2017 18:01	SMPL
I1133-11	ME5BM5	6912	1.5298	-	ISM02.3	PPB	1/18/2017 18:03	SMPL
I1133-12	ME5BM6	3175	0.7121	-	ISM02.3	PPB	1/18/2017 18:06	SMPL
I1133-13	ME5BM7	2233	0.5059	-	ISM02.3	PPB	1/18/2017 18:08	SMPL
I1133-14	ME5BM8	2128	0.483	-	ISM02.3	PPB	1/18/2017 18:10	SMPL
I1133-15	ME5BM9	1902	0.4335	-	ISM02.3	PPB	1/18/2017 18:12	SMPL
I1133-16	ME5BN0	2859	0.6429	-	ISM02.3	PPB	1/18/2017 18:14	SMPL
I1133-17	ME5BN1	18230	4.0064	-	ISM02.3	PPB	1/18/2017 18:16	SMPL
I1133-18	ME5BN2	9140	2.0173	-	ISM02.3	PPB	1/18/2017 18:18	SMPL
I1133-19	ME5BN3	56413	12.3616	-	ISM02.3	PPB	1/18/2017 18:20	SMPL
CCV080	CCV080	23524	5.1648	-	ISM02.3	PPB	1/18/2017 18:23	SMPL
CCB080	CCB080	-20	0.0129	-	ISM02.3	PPB	1/18/2017 18:25	SMPL
I1133-20	ME5BN4	9464	2.0882	-	ISM02.3	PPB	1/18/2017 18:27	SMPL
I1133-21	ME5BN4D	9131	2.0153	-	ISM02.3	PPB	1/18/2017 18:29	SMPL
I1133-22	ME5BN4S	20513	4.5059	-	ISM02.3	PPB	1/18/2017 18:31	SMPL
PB96075BL	PBS004	-36	0.0094	-	ISM02.3	PPB	1/18/2017 18:33	SMPL
I1226-01	ME5BN5	1392	0.3219	-	ISM02.3	PPB	1/18/2017 18:35	SMPL
I1226-02	ME5BN6	4226	0.942	-	ISM02.3	PPB	1/18/2017 18:37	SMPL
I1226-03	ME5BN7	4496	1.0011	-	ISM02.3	PPB	1/18/2017 18:40	SMPL
I1226-04	ME5BN8	4236	0.9442	-	ISM02.3	PPB	1/18/2017 18:42	SMPL
I1226-05	ME5BN9	6474	1.4339	-	ISM02.3	PPB	1/18/2017 18:44	SMPL
I1226-06	ME5BP0	4557	1.0145	-	ISM02.3	PPB	1/18/2017 18:46	SMPL
I1226-07	ME5BP1	10869	2.3957	-	ISM02.3	PPB	1/18/2017 18:48	SMPL
I1226-08	ME5BP2	9370	2.0676	-	ISM02.3	PPB	1/18/2017 18:50	SMPL
I1226-09	ME5BP3	19758	4.3407	-	ISM02.3	PPB	1/18/2017 18:52	SMPL
I1226-10	ME5BP4	5765	1.2788	-	ISM02.3	PPB	1/18/2017 18:54	SMPL

INSTRUMENT ID : CV1

I1226-11	ME5BP5	5446	1.209 -	ISM02.3	PPB	1/18/2017 18:57 SMPL
I1226-12	ME5BP6	3154	0.7075 -	ISM02.3	PPB	1/18/2017 18:59 SMPL
I1226-13	ME5BP7	6712	1.486 -	ISM02.3	PPB	1/18/2017 19:01 SMPL
I1226-14	ME5BP8	2433	0.5497 -	ISM02.3	PPB	1/18/2017 19:03 SMPL
I1226-15	ME5BP9	2744	0.6177 -	ISM02.3	PPB	1/18/2017 19:05 SMPL
I1226-16	ME5BQ0	17010	3.7394 -	ISM02.3	PPB	1/18/2017 19:07 SMPL
CCV081	CCV081	23637	5.1895 -	ISM02.3	PPB	1/18/2017 19:09 SMPL
CCB081	CCB081	-11	0.0149 -	ISM02.3	PPB	1/18/2017 19:11 SMPL
I1226-17	ME5BQ1	25075	5.5042 -	ISM02.3	PPB	1/18/2017 19:14 SMPL
I1226-18	ME5BQ2	27325	5.9965 -	ISM02.3	PPB	1/18/2017 19:16 SMPL
I1226-19	ME5BQ3	34375	7.5392 -	ISM02.3	PPB	1/18/2017 19:18 SMPL
I1226-20	ME5BQ4	9204	2.0313 -	ISM02.3	PPB	1/18/2017 19:20 SMPL
I1226-21	ME5BQ4D	8272	1.8274 -	ISM02.3	PPB	1/18/2017 19:22 SMPL
I1226-22	ME5BQ4S	20541	4.5121 -	ISM02.3	PPB	1/18/2017 19:24 SMPL
I1133-19DLX5	ME5BN3	11216	2.4716 -	ISM02.3	PPB	1/18/2017 19:26 SMPL
CCV082	CCV082	23722	5.2081 -	ISM02.3	PPB	1/18/2017 19:28 SMPL
CCB082	CCB082	-7	0.0158 -	ISM02.3	PPB	1/18/2017 19:31 SMPL