



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB85406
Balance ID #: Gravimetric
Extraction IN Time: 01/19/2017 08:15

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 01/19/2017
Analysis Date: 01/19/2017
Extraction OUT Time: 01/19/2017 09:00

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight	Weight After Drying	Final Weight After Drying	Change Weight (g)	Result in ppm
1	LB85406BL	LB85406BL	WATER	2	1000	100	2.4136	2.4136	2.4136	2.4136	0	0
2	LB85406BS	LB85406BS	WATER	2	1000	100	2.9652	2.9652	2.9842	2.9842	0.019	19
3	LB85406BSD	LB85406BSD	WATER	2	1000	100	2.9394	2.9394	2.9586	2.9586	0.0192	19.2
4	I1306-01	STORMDRAIN-1	WATER	2	970	100	2.9504	2.9504	2.9506	2.9506	0.0002	0.21
5	I1306-01MS	STORMDRAIN-1	WATER	2	980	100	2.9714	2.9714	2.9897	2.9897	0.0183	18.67
6	I1306-02	STORMDRAIN-2	WATER	2	970	100	2.9171	2.9171	2.9176	2.9176	0.0005	0.52
7	I1306-02DUP	STORMDRAIN-2DUP	WATER	2	970	100	2.9384	2.9384	2.939	2.939	0.0006	0.62



QC Batch# LB85406 Test: TPH Analysis Date: 01/19/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2196
Magnesium Sulfate	
Sodium Sulfate	EP1722
1:1 HCL	WP50630
Silica Gel	W2061 (3g)
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP50631
LCSWD	10.00 ML	WP50631
MS/MSD	10.00 ML	WP50631

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID #

Before Analysis

0.0020 gram Balance: 0.0019 In OVEN TEMP : 70 Dessicator Time In : 10:00
1.0000 gram Balance: 1.0005 In Time: 09:30
Bal Check Time: 08:30

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 11:30
1.0000 gram Balance: 1.0003 Out Time: 11:00
Bal Check Time: 12:30

1385406

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH I1306

WorkList ID : 94572

Date : 1/19/2017 8:12:44 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1306-01	TPH	Conc H2SO4 4c	KWAR01	D51	STORMDRAIN-1	01/18/2017	1664A
	Water	I1306-02	TPH	Conc H2SO4 4c	KWAR01	D51	STORMDRAIN-2	01/18/2017	1664A

Date/Time 01-19-17 8:00 AM
Received by: SL
Relinquished by: CS

Date/Time 01-19-17 10:20 AM
Received by: CS
Relinquished by: JP