


7470A

LB 85409
INSTRUMENT ID: CV1

Linear ▼

 μ Abs.:

39146



A= 0.0000e+000

B= 2.5801e-004

C= 3.0126e-002

Rho= 0.9996909

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.032	0.032	9	0.000	9				
0.2	0.200	0.224	0.024	750	0.0 %	750				
2.5	2.500	2.492	-0.008	9541	0.0 %	9541				
5.0	5.000	5.000	0.000	19264	0.0 %	19264				
7.5	7.500	7.321	-0.179	28260	0.0 %	28260				
10.0	10.000	10.130	0.130	39146	0.0 %	39146				

LB85409 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Jan 2017 10:36:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	9	-	0.00007470A	PPB		19 Jan 2017 10:40:40	S	Std
0.2 - 1	S0.2	750	-	0.20007470A	PPB		19 Jan 2017 10:42:49	S	Std
2.5 - 1	S2.5	9541	-	2.50007470A	PPB		19 Jan 2017 10:44:57	S	Std
5.0 - 1	S5	19264	-	5.00007470A	PPB		19 Jan 2017 10:47:05	S	Std
7.5 - 1	S7.5	28260	-	7.50007470A	PPB		19 Jan 2017 10:49:13	S	Std
10.0 - 1	S10	39146	-	10.00007470A	PPB		19 Jan 2017 10:51:20	S	Std
ICV03 - 1	ICV03	15250	3.9648	-7470A	PPB		19 Jan 2017 10:56:19	U	SMPL
ICB03 - 1	ICB03	-32	0.0219	-7470A	PPB		19 Jan 2017 10:58:24	U	SMPL
CCV83 - 1	CCV83	19435	5.0446	-7470A	PPB		19 Jan 2017 11:00:31	U	SMPL
CCB83 - 1	CCB83	-28	0.0229	-7470A	PPB		19 Jan 2017 11:02:37	U	SMPL
CRA - 1	CRA	753	0.2244	-7470A	PPB		19 Jan 2017 11:04:44	U	SMPL
HighStd - 1	HighStd	38761	10.0309	-7470A	PPB		19 Jan 2017 11:06:50	U	SMPL
ChkStd - 1	ChkStd	26923	6.9765	-7470A	PPB		19 Jan 2017 11:08:57	U	SMPL
PB96103BL - 1	PBW	-16	0.0260	-7470A	PPB		19 Jan 2017 11:11:04	U	SMPL
PB96103BS - 1	LCSW	14416	3.7496	-7470A	PPB		19 Jan 2017 11:13:12	U	SMPL
PB96103TB - 1	PB96078TB	7	0.0319	-7470A	PPB		19 Jan 2017 11:15:19	U	SMPL
I1260-04 - 1	MH-0-12FT-COMP	22	0.0358	-7470A	PPB		19 Jan 2017 11:17:27	U	SMPL
I1260-04DUP - 1	MH-0-12FT-COMPDUP	6	0.0317	-7470A	PPB		19 Jan 2017 11:19:36	U	SMPL
I1260-04MS - 1	MH-0-12FT-COMPMS	16225	4.2163	-7470A	PPB		19 Jan 2017 11:21:45	U	SMPL
I1260-04MSD - 1	MH-0-12FT-COMPMSD	16321	4.2411	-7470A	PPB		19 Jan 2017 11:26:26	U	SMPL
CCV84 - 1	CCV84	19411	5.0384	-7470A	PPB		19 Jan 2017 11:28:32	U	SMPL
CCB84 - 1	CCB84	-28	0.0229	-7470A	PPB		19 Jan 2017 11:30:39	U	SMPL
PB96122BL - 1	PBW	4	0.0312	-7470A	PPB		19 Jan 2017 11:32:46	U	SMPL
PB96122BS - 1	LCSW	14266	3.7109	-7470A	PPB		19 Jan 2017 11:34:53	U	SMPL
I1265-02 - 1	B-1WS	1904	0.5214	-7470A	PPB		19 Jan 2017 11:36:59	U	SMPL
I1265-05 - 1	B-3WS	1409	0.3937	-7470A	PPB		19 Jan 2017 11:39:07	U	SMPL
I1265-07 - 1	B-4WS	237	0.0913	-7470A	PPB		19 Jan 2017 11:41:14	U	SMPL
I1265-07DUP - 1	B-4WSDUP	259	0.0970	-7470A	PPB		19 Jan 2017 11:43:22	U	SMPL
I1265-07MS - 1	B-4WSMS	16112	4.1872	-7470A	PPB		19 Jan 2017 11:45:30	U	SMPL
I1265-07MSD - 1	B-4WSMSD	17175	4.4615	-7470A	PPB		19 Jan 2017 11:47:38	U	SMPL
CCV85 - 1	CCV85	19208	4.9860	-7470A	PPB		19 Jan 2017 11:49:47	U	SMPL
CCB85 - 1	CCB85	-15	0.0263	-7470A	PPB		19 Jan 2017 11:51:54	U	SMPL
I1260-04LX5 - 1		23	0.0361	-7470A	PPB		19 Jan 2017 11:54:01	U	SMPL
I1260-04A - 1		15520	4.0344	-7470A	PPB		19 Jan 2017 11:56:08	U	SMPL
I1265-07LX5 - 1		45	0.0417	-7470A	PPB		19 Jan 2017 11:58:15	U	SMPL
I1265-07A - 1		17308	4.4958	-7470A	PPB		19 Jan 2017 12:00:22	U	SMPL
CCV86 - 1	CCV86	19467	5.0528	-7470A	PPB		19 Jan 2017 12:02:29	U	SMPL
CCB86 - 1	CCB86	-31	0.0221	-7470A	PPB		19 Jan 2017 12:04:36	U	SMPL