


7471A

LB85413
INSTRUMENT ID: CV8

Linear

 μ Abs.:

46313



A= 0.0000e+000

B= 2.1588e-004

C= 8.5766e-003

Rho= 0.9999897

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	27	0.000	27				
0.2	0.200	0.201	0.001	892	0.0 %	892				
2.5	2.500	2.470	-0.030	11404	0.0 %	11404				
5.0	5.000	5.020	0.020	23213	0.0 %	23213				
7.5	7.500	7.487	-0.013	34643	0.0 %	34643				
10.0	10.000	10.007	0.007	46313	0.0 %	46313				

LB85413 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 19 Jan 2017 12:28:54

Sample ID	Extended ID	μ Abs.	Conc.	Stdnd Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	27	-	0.0000	7471A	PPB	19 Jan 2017 12:31:44	S	Std
0.2 - 1	S0.2	892	-	0.2000	7471A	PPB	19 Jan 2017 12:34:02	S	Std
2.5 - 1	S2.5	11404	-	2.5000	7471A	PPB	19 Jan 2017 12:36:20	S	Std
5.0 - 1	S5	23213	-	5.0000	7471A	PPB	19 Jan 2017 12:38:37	S	Std
7.5 - 1	S7.5	34643	-	7.5000	7471A	PPB	19 Jan 2017 12:40:54	S	Std
10.0 - 1	S10	46313	-	10.0000	7471A	PPB	19 Jan 2017 12:43:11	S	Std
ICV04 - 1	ICV04	17841	3.8601	-	7471A	PPB	19 Jan 2017 12:45:58	U	SMPL
ICB04 - 1	ICB04	-12	0.0060	-	7471A	PPB	19 Jan 2017 12:48:14	U	SMPL
CCV87 - 1	CCV87	23154	5.0071	-	7471A	PPB	19 Jan 2017 12:50:29	U	SMPL
CCB87 - 1	CCB87	-16	0.0051	-	7471A	PPB	19 Jan 2017 12:52:45	U	SMPL
CRA - 1	CRA	881	0.1988	-	7471A	PPB	19 Jan 2017 12:55:00	U	SMPL
HighStd - 1	HighStd	45572	9.8468	-	7471A	PPB	19 Jan 2017 12:57:16	U	SMPL
ChkStd - 1	ChkStd	31930	6.9017	-	7471A	PPB	19 Jan 2017 12:59:32	U	SMPL
PB96124BL - 1	PBS	-42	-0.0005	-	7471A	PPB	19 Jan 2017 13:01:48	U	SMPL
PB96124BS - 1	LCSS	14836	3.2114	-	7471A	PPB	19 Jan 2017 13:04:05	U	SMPL
I1260-01 - 1	MH-0-12FT-COMP	1468	0.3255	-	7471A	PPB	19 Jan 2017 13:06:22	U	SMPL
I1260-01DUP - 1	MH-0-12FT-COMPDUP	1559	0.3451	-	7471A	PPB	19 Jan 2017 13:08:39	U	SMPL
I1260-01MS - 1	MH-0-12FT-COMPMS	17859	3.8640	-	7471A	PPB	19 Jan 2017 13:10:57	U	SMPL
I1260-01MSD - 1	MH-0-12FT-COMPMSD	18902	4.0892	-	7471A	PPB	19 Jan 2017 13:13:15	U	SMPL
I1265-01 - 1	B-1	4199	0.9151	-	7471A	PPB	19 Jan 2017 13:15:31	U	SMPL
CCV88 - 1	CCV88	23151	5.0065	-	7471A	PPB	19 Jan 2017 13:17:46	U	SMPL
CCB88 - 1	CCB88	-22	0.0038	-	7471A	PPB	19 Jan 2017 13:20:02	U	SMPL
I1265-03 - 1	B-2	6472	1.4058	-	7471A	PPB	19 Jan 2017 13:22:18	U	SMPL
I1265-04 - 1	B-3	5806	1.2620	-	7471A	PPB	19 Jan 2017 13:24:34	U	SMPL
I1265-06 - 1	B-4	1573	0.3482	-	7471A	PPB	19 Jan 2017 13:26:51	U	SMPL
I1265-08 - 1	B-5	4961	1.0796	-	7471A	PPB	19 Jan 2017 13:29:07	U	SMPL
I1265-09 - 1	B-6	2851	0.6241	-	7471A	PPB	19 Jan 2017 13:31:24	U	SMPL
I1271-02 - 1	SS-01	7482	1.6238	-	7471A	PPB	19 Jan 2017 13:33:41	U	SMPL
I1271-03 - 1	SS-02	2012	0.4429	-	7471A	PPB	19 Jan 2017 13:35:58	U	SMPL
I1260-01LX5 - 1		285	0.0701	-	7471A	PPB	19 Jan 2017 13:38:16	U	SMPL
I1260-01A - 1		19237	4.1615	-	7471A	PPB	19 Jan 2017 13:40:34	U	SMPL
CCV89 - 1	CCV89	22775	4.9253	-	7471A	PPB	19 Jan 2017 13:42:50	U	SMPL
CCB89 - 1	CCB89	-20	0.0043	-	7471A	PPB	19 Jan 2017 13:45:06	U	SMPL