



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85415

Temp.: 22.6 °C

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/19/2017	11:00
2	CAL2	1	Water			7.02	01/19/2017	11:01
3	CAL3	1	Water			10.04	01/19/2017	11:02
4	ICV	1	Water			7.01	01/19/2017	11:03
5	CCV1	1	Water			2.02	01/19/2017	11:04
6	I1304-03	1	Solid	20.02	20	6.82	01/19/2017	11:05
7	I1304-03DUP	1	Solid	20.03	20	6.83	01/19/2017	11:06
8	I1304-06	1	Solid	20.04	20	6.64	01/19/2017	11:07
9	I1304-09	1	Solid	20.05	20	7.68	01/19/2017	11:09
10	I1308-04	1	Solid	20.01	20	6.89	01/19/2017	11:10
11	CCV2	1	Water			12.02	01/19/2017	11:11

2385415

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I1304			WorkList ID : 94601		Date : 1/19/2017 10:47:48 AM				
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/23/2017	Solid	I1304-03	Corrosivity	Cool 4 deg C	PSEG01	D52	ROLL-OFF-15-41-COMP	01/18/2017	9045C
01/23/2017	Solid	I1304-06	Corrosivity	Cool 4 deg C	PSEG01	D52	1634-NOTTINGHAM-COMP	01/18/2017	9045C
01/23/2017	Solid	I1304-09	Corrosivity	Cool 4 deg C	PSEG01	D52	ROLL-OFF-15939-COMP	01/18/2017	9045C
01/23/2017	Solid	I1308-04	Corrosivity	Cool 4 deg C	PSEG01	D52	MH-2-0-12FT-COMP	01/18/2017	9045C

Date/Time 01-19-17 14:40:07
Received by: JN
Relinquished by: JN

Date/Time 01-19-17 12:30 PM
Received by: CD
Relinquished by: JN