



Analytical Summary Report

Analysis Method: 9045C
Parameter: pH
Run Number: LB85416

Analyst By : JIGNESH
Supervisor Review By : apatel
Temp.: 16.8 °C
SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	01/19/2017	13:00
2	CAL2	1	Water			7.01	01/19/2017	13:01
3	CAL3	1	Water			10.03	01/19/2017	13:02
4	ICV	1	Water			7.02	01/19/2017	13:03
5	CCV1	1	Water			2.01	01/19/2017	13:04
6	I1312-13	1	Water			8.86	01/19/2017	13:05
7	I1312-13DUP	1	Water			8.87	01/19/2017	13:06
8	CCV2	1	Water			12.01	01/19/2017	13:07

1385416

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I1312

WorkList ID : 94613

Date : 1/19/2017 12:44:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1312-13	pH	Cool 4 deg C	PARS15	D11	CAS-WC011817-LIQUID	01/18/2017	9045C

Date/Time 01-19-17 12:44pm
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 01-19-17 21:00 [Signature]
Received by: [Signature]
Relinquished by: [Signature]