


ISM02.3

LB 85417
INSTRUMENT ID: CV1 μ Abs.:

41447



Linear

A= 0.0000e+000

B= 2.4244e-004

C= -4.0662e-002

Rho= 0.9999672

Accept= Accepted

Slope
Intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.017	-0.017	96	0.000	96					-7
0.20	0.200	0.187	-0.013	937	0.0 %	937					1
2.50	2.500	2.531	0.031	10606	0.0 %	10606					1
5.00	5.000	5.040	0.040	20957	0.0 %	20957					-1
7.50	7.500	7.452	-0.048	30907	0.0 %	30907					0
10.00	10.000	10.008	0.008	41447	0.0 %	41447					

NB
1/19/17

LB85417
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	96	-		0 ISM02.3	PPB	1/19/2017 14:27	Std
	0.2 S0.2	937	-		0.2 ISM02.3	PPB	1/19/2017 14:29	Std
	2.5 S2.5	10606	-		2.5 ISM02.3	PPB	1/19/2017 14:31	Std
	5 S5	20957	-		5 ISM02.3	PPB	1/19/2017 14:34	Std
	7.5 S7.5	30907	-		7.5 ISM02.3	PPB	1/19/2017 14:36	Std
	10 S10	41447	-		10 ISM02.3	PPB	1/19/2017 14:38	Std
ICV	ICV	16560	3.9741 -		ISM02.3	PPB	1/19/2017 14:41	SMPL
ICB	ICB	55	-0.0273 -		ISM02.3	PPB	1/19/2017 14:43	SMPL
CCV090	CCV090	20955	5.0396 -		ISM02.3	PPB	1/19/2017 14:45	SMPL
CCB090	CCB090	57	-0.0268 -		ISM02.3	PPB	1/19/2017 14:47	SMPL
PB96097BL	PBW003	7	-0.039 -		ISM02.3	PPB	1/19/2017 14:49	SMPL
I1252-01	MC0CF3	23	-0.0351 -		ISM02.3	PPB	1/19/2017 14:51	SMPL
I1252-02	MC0CG3	-3	-0.0414 -		ISM02.3	PPB	1/19/2017 14:53	SMPL
I1252-03	MC0CL8	-3	-0.0414 -		ISM02.3	PPB	1/19/2017 14:56	SMPL
I1252-04	MC0CL8D	0	-0.0407 -		ISM02.3	PPB	1/19/2017 14:58	SMPL
I1252-05	MC0CL8S	4419	1.0307 -		ISM02.3	PPB	1/19/2017 15:00	SMPL
I1252-06	MC0CP1	-9	-0.0428 -		ISM02.3	PPB	1/19/2017 15:02	SMPL
I1252-07	MC0CF9	-10	-0.0431 -		ISM02.3	PPB	1/19/2017 15:04	SMPL
I1252-08	MC0CH9	252	0.0204 -		ISM02.3	PPB	1/19/2017 15:06	SMPL
I1252-09	MC0CL4	-1	-0.0409 -		ISM02.3	PPB	1/19/2017 15:08	SMPL
I1252-10	MC0CQ4	21	-0.0356 -		ISM02.3	PPB	1/19/2017 15:10	SMPL
I1252-11	MC0CQ6	-6	-0.0421 -		ISM02.3	PPB	1/19/2017 15:12	SMPL
I1252-12	MC0CR7	-2	-0.0411 -		ISM02.3	PPB	1/19/2017 15:15	SMPL
I1252-13	MC0CG1	0	-0.0407 -		ISM02.3	PPB	1/19/2017 15:17	SMPL
I1252-14	MC0CG9	7	-0.039 -		ISM02.3	PPB	1/19/2017 15:19	SMPL
I1252-15	MC0CJ9	1	-0.0404 -		ISM02.3	PPB	1/19/2017 15:21	SMPL
I1252-16	MC0CK1	21	-0.0356 -		ISM02.3	PPB	1/19/2017 15:23	SMPL
I1252-17	MC0CQ8	-9	-0.0428 -		ISM02.3	PPB	1/19/2017 15:25	SMPL
I1252-18	MC0CH5	14	-0.0373 -		ISM02.3	PPB	1/19/2017 15:27	SMPL
I1252-19	MC0CJ5	15	-0.037 -		ISM02.3	PPB	1/19/2017 15:29	SMPL
CCV091	CCV091	20916	5.0302 -		ISM02.3	PPB	1/19/2017 15:32	SMPL
CCB091	CCB091	58	-0.0266 -		ISM02.3	PPB	1/19/2017 15:34	SMPL
I1252-20	MC0CK4	-6	-0.0421 -		ISM02.3	PPB	1/19/2017 15:36	SMPL
I1252-21	MC0CK6	-3	-0.0414 -		ISM02.3	PPB	1/19/2017 15:38	SMPL
I1252-22	MC0CQ1	7	-0.039 -		ISM02.3	PPB	1/19/2017 15:40	SMPL
PB96098BL	PBW003	4	-0.0397 -		ISM02.3	PPB	1/19/2017 15:42	SMPL
I1254-01	MC0CF2	0	-0.0407 -		ISM02.3	PPB	1/19/2017 15:44	SMPL
I1254-02	MC0CG2	-5	-0.0419 -		ISM02.3	PPB	1/19/2017 15:46	SMPL
I1254-03	MC0CL9	-5	-0.0419 -		ISM02.3	PPB	1/19/2017 15:49	SMPL
I1254-04	MC0CL9D	-5	-0.0419 -		ISM02.3	PPB	1/19/2017 15:51	SMPL
I1254-05	MC0CL9S	4345	1.0127 -		ISM02.3	PPB	1/19/2017 15:53	SMPL
I1254-06	MC0CP2	-10	-0.0431 -		ISM02.3	PPB	1/19/2017 15:55	SMPL
I1254-07	MC0CF8	-6	-0.0421 -		ISM02.3	PPB	1/19/2017 15:57	SMPL
I1254-08	MC0CH8	-1	-0.0409 -		ISM02.3	PPB	1/19/2017 15:59	SMPL
I1254-09	MC0CL5	10	-0.0382 -		ISM02.3	PPB	1/19/2017 16:01	SMPL
I1254-10	MC0CQ5	2	-0.0402 -		ISM02.3	PPB	1/19/2017 16:03	SMPL

I1254-11	MC0CQ7	0	-0.0407 -	ISM02.3	PPB	1/19/2017 16:06	SMPL
I1254-12	MC0CG0	-12	-0.0436 -	ISM02.3	PPB	1/19/2017 16:08	SMPL
I1254-13	MC0CG8	-2	-0.0411 -	ISM02.3	PPB	1/19/2017 16:10	SMPL
I1254-14	MC0CK0	-10	-0.0431 -	ISM02.3	PPB	1/19/2017 16:12	SMPL
I1254-15	MC0CK2	-14	-0.0441 -	ISM02.3	PPB	1/19/2017 16:14	SMPL
I1254-16	MC0CQ9	-8	-0.0426 -	ISM02.3	PPB	1/19/2017 16:16	SMPL
CCV092	CCV092	20992	5.0486 -	ISM02.3	PPB	1/19/2017 16:18	SMPL
CCB092	CCB092	51	-0.0283 -	ISM02.3	PPB	1/19/2017 16:20	SMPL
I1254-17	MC0CH4	-19	-0.0453 -	ISM02.3	PPB	1/19/2017 16:23	SMPL
I1254-18	MC0CJ4	-13	-0.0438 -	ISM02.3	PPB	1/19/2017 16:25	SMPL
I1254-19	MC0CK5	-14	-0.0441 -	ISM02.3	PPB	1/19/2017 16:27	SMPL
I1254-20	MC0CK7	-16	-0.0445 -	ISM02.3	PPB	1/19/2017 16:29	SMPL
PB96099BL	PBW003	-1	-0.0409 -	ISM02.3	PPB	1/19/2017 16:31	SMPL
I1255-01	MC0CE7	0	-0.0407 -	ISM02.3	PPB	1/19/2017 16:33	SMPL
I1255-02	MC0CF5	46	-0.0295 -	ISM02.3	PPB	1/19/2017 16:35	SMPL
I1255-03	MC0CF7	-13	-0.0438 -	ISM02.3	PPB	1/19/2017 16:37	SMPL
I1255-04	MC0CJ1	519	0.0852 -	ISM02.3	PPB	1/19/2017 16:39	SMPL
I1255-05	MC0CP8	-3	-0.0414 -	ISM02.3	PPB	1/19/2017 16:42	SMPL
I1255-06	MC0CG5	2	-0.0402 -	ISM02.3	PPB	1/19/2017 16:44	SMPL
I1255-07	MC0CG7	15	-0.037 -	ISM02.3	PPB	1/19/2017 16:46	SMPL
I1255-08	MC0CH1	5	-0.0394 -	ISM02.3	PPB	1/19/2017 16:48	SMPL
I1255-09	MC0CJ7	-11	-0.0433 -	ISM02.3	PPB	1/19/2017 16:50	SMPL
I1255-10	MC0CJ7D	-7	-0.0424 -	ISM02.3	PPB	1/19/2017 16:52	SMPL
I1255-11	MC0CJ7S	4341	1.0118 -	ISM02.3	PPB	1/19/2017 16:54	SMPL
I1255-12	MC0CK8	-15	-0.0443 -	ISM02.3	PPB	1/19/2017 16:56	SMPL
I1255-13	MC0CM2	8	-0.0387 -	ISM02.3	PPB	1/19/2017 16:59	SMPL
I1255-14	MC0CM4	32	-0.0329 -	ISM02.3	PPB	1/19/2017 17:01	SMPL
I1255-15	MC0CM6	35	-0.0322 -	ISM02.3	PPB	1/19/2017 17:03	SMPL
CCV093	CCV093	21056	5.0641 -	ISM02.3	PPB	1/19/2017 17:05	SMPL
CCB093	CCB093	111	-0.0138 -	ISM02.3	PPB	1/19/2017 17:07	SMPL
I1255-16	MC0CP6	0	-0.0407 -	ISM02.3	PPB	1/19/2017 17:09	SMPL
I1255-17	MC0CR0	-2	-0.0411 -	ISM02.3	PPB	1/19/2017 17:11	SMPL
I1255-18	MC0CL6	86	-0.0198 -	ISM02.3	PPB	1/19/2017 17:13	SMPL
I1255-19	MC0CM0	85	-0.0201 -	ISM02.3	PPB	1/19/2017 17:16	SMPL
I1255-20	MC0CN6	528	0.0873 -	ISM02.3	PPB	1/19/2017 17:18	SMPL
I1255-21	MC0CN7	33	-0.0327 -	ISM02.3	PPB	1/19/2017 17:20	SMPL
I1255-22	MC0CN9	7	-0.039 -	ISM02.3	PPB	1/19/2017 17:22	SMPL
PB96117BL	PBW003	3	-0.0399 -	ISM02.3	PPB	1/19/2017 17:24	SMPL
I1256-01	MC0CE6	1	-0.0404 -	ISM02.3	PPB	1/19/2017 17:26	SMPL
I1256-02	MC0CF4	-4	-0.0416 -	ISM02.3	PPB	1/19/2017 17:28	SMPL
I1256-03	MC0CF6	-2	-0.0411 -	ISM02.3	PPB	1/19/2017 17:31	SMPL
I1256-04	MC0CJ0	-1	-0.0409 -	ISM02.3	PPB	1/19/2017 17:33	SMPL
I1256-05	MC0CK3	-7	-0.0424 -	ISM02.3	PPB	1/19/2017 17:35	SMPL
I1256-06	MC0CP9	4	-0.0397 -	ISM02.3	PPB	1/19/2017 17:37	SMPL
I1256-07	MC0CG4	3	-0.0399 -	ISM02.3	PPB	1/19/2017 17:39	SMPL
I1256-08	MC0CG6	-19	-0.0453 -	ISM02.3	PPB	1/19/2017 17:41	SMPL
I1256-09	MC0CHO	-5	-0.0419 -	ISM02.3	PPB	1/19/2017 17:43	SMPL

LB85417
INSTRUMENT ID : CV1

I1256-10	MCOCJ6	-11	-0.0433 -	ISM02.3	PPB	1/19/2017 17:45	SMPL
I1256-11	MCOCJ6D	-10	-0.0431 -	ISM02.3	PPB	1/19/2017 17:48	SMPL
I1256-12	MCOCJ6S	4356	1.0154 -	ISM02.3	PPB	1/19/2017 17:50	SMPL
CCV094	CCV094	21024	5.0564 -	ISM02.3	PPB	1/19/2017 17:52	SMPL
CCB094	CCB094	54	-0.0276 -	ISM02.3	PPB	1/19/2017 17:54	SMPL
I1256-13	MCOCJ8	-19	-0.0453 -	ISM02.3	PPB	1/19/2017 17:56	SMPL
I1256-14	MCOCK9	-12	-0.0436 -	ISM02.3	PPB	1/19/2017 17:58	SMPL
I1256-15	MCOCM3	-9	-0.0428 -	ISM02.3	PPB	1/19/2017 18:00	SMPL
I1256-16	MCOCM5	-4	-0.0416 -	ISM02.3	PPB	1/19/2017 18:02	SMPL
I1256-17	MCOCM7	0	-0.0407 -	ISM02.3	PPB	1/19/2017 18:05	SMPL
I1256-18	MCOCL7	7	-0.039 -	ISM02.3	PPB	1/19/2017 18:07	SMPL
I1256-19	MCOCM1	7	-0.039 -	ISM02.3	PPB	1/19/2017 18:09	SMPL
I1256-20	MCOCN5	23	-0.0351 -	ISM02.3	PPB	1/19/2017 18:11	SMPL
I1256-21	MCOCN8	-1	-0.0409 -	ISM02.3	PPB	1/19/2017 18:13	SMPL
I1256-22	MCOCPO	-4	-0.0416 -	ISM02.3	PPB	1/19/2017 18:15	SMPL
PB96118BL	PBW003	1	-0.0404 -	ISM02.3	PPB	1/19/2017 18:17	SMPL
I1257-01	MCOCCE4	-10	-0.0431 -	ISM02.3	PPB	1/19/2017 18:19	SMPL
I1257-02	MCOCF1	6	-0.0392 -	ISM02.3	PPB	1/19/2017 18:22	SMPL
I1257-03	MCOCM8	46	-0.0295 -	ISM02.3	PPB	1/19/2017 18:24	SMPL
I1257-04	MCOCNO	1601	0.3475 -	ISM02.3	PPB	1/19/2017 18:26	SMPL
I1257-05	MCOCN2	-5	-0.0419 -	ISM02.3	PPB	1/19/2017 18:28	SMPL
I1257-06	MCOCR2	-2	-0.0411 -	ISM02.3	PPB	1/19/2017 18:30	SMPL
I1257-07	MCOCCE9	-5	-0.0419 -	ISM02.3	PPB	1/19/2017 18:32	SMPL
I1257-08	MCOCH3	-1	-0.0409 -	ISM02.3	PPB	1/19/2017 18:34	SMPL
I1257-09	MCOCH3D	0	-0.0407 -	ISM02.3	PPB	1/19/2017 18:37	SMPL
CCV095	CCV095	20764	4.9933 -	ISM02.3	PPB	1/19/2017 18:39	SMPL
CCB095	CCB095	63	-0.0254 -	ISM02.3	PPB	1/19/2017 18:41	SMPL
I1257-10	MCOCH3S	4595	1.0733 -	ISM02.3	PPB	1/19/2017 18:43	SMPL
I1257-11	MCOCPC4	-9	-0.0428 -	ISM02.3	PPB	1/19/2017 18:45	SMPL
I1257-12	MCOCQ2	22	-0.0353 -	ISM02.3	PPB	1/19/2017 18:47	SMPL
I1257-13	MCOCR4	379	0.0512 -	ISM02.3	PPB	1/19/2017 18:49	SMPL
I1257-14	MCOCR6	-6	-0.0421 -	ISM02.3	PPB	1/19/2017 18:51	SMPL
PB96123BL	PBW003	57	-0.0268 -	ISM02.3	PPB	1/19/2017 18:54	SMPL
I1258-01	MCOCCE5	57	-0.0268 -	ISM02.3	PPB	1/19/2017 18:56	SMPL
I1258-02	MCOCCE8	48	-0.029 -	ISM02.3	PPB	1/19/2017 18:58	SMPL
I1258-03	MCOCM9	60	-0.0261 -	ISM02.3	PPB	1/19/2017 19:00	SMPL
I1258-04	MCOCN1	46	-0.0295 -	ISM02.3	PPB	1/19/2017 19:02	SMPL
I1258-05	MCOCN3	65	-0.0249 -	ISM02.3	PPB	1/19/2017 19:04	SMPL
I1258-06	MCOCFO	51	-0.0283 -	ISM02.3	PPB	1/19/2017 19:06	SMPL
I1258-07	MCOCH2	63	-0.0254 -	ISM02.3	PPB	1/19/2017 19:09	SMPL
I1258-08	MCOCH2D	67	-0.0244 -	ISM02.3	PPB	1/19/2017 19:11	SMPL
I1258-09	MCOCH2S	4316	1.0057 -	ISM02.3	PPB	1/19/2017 19:13	SMPL
I1258-10	MCOCPC5	61	-0.0259 -	ISM02.3	PPB	1/19/2017 19:15	SMPL
I1258-11	MCOCQ3	75	-0.0225 -	ISM02.3	PPB	1/19/2017 19:17	SMPL
I1258-12	MCOCR8	82	-0.0208 -	ISM02.3	PPB	1/19/2017 19:19	SMPL
I1258-13	MCOCR1	66	-0.0247 -	ISM02.3	PPB	1/19/2017 19:21	SMPL
I1258-14	MCOCR5	85	-0.0201 -	ISM02.3	PPB	1/19/2017 19:24	SMPL

LB85417

INSTRUMENT ID : CV1

CCV096	CCV096	20887	5.0232 -	ISM02.3	PPB	1/19/2017 19:26 SMPL
CCB096	CCB096	66	-0.0247 -	ISM02.3	PPB	1/19/2017 19:28 SMPL
I1258-15	MCOCR3	78	-0.0218 -	ISM02.3	PPB	1/19/2017 19:30 SMPL
CCV097	CCV097	20933	5.0343 -	ISM02.3	PPB	1/19/2017 19:32 SMPL
CCB097	CCB097	60	-0.0261 -	ISM02.3	PPB	1/19/2017 19:34 SMPL