



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85430

Temp.: 23.2 °C

BalanceID:

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	01/20/2017	10:45
2	CAL2	1	Water			7.01	01/20/2017	10:46
3	CAL3	1	Water			10.02	01/20/2017	10:47
4	ICV	1	Water			7.01	01/20/2017	10:48
5	CCV1	1	Water			2.01	01/20/2017	10:49
6	I1347-02	1	Solid	20.02	20	8.25	01/20/2017	10:50
7	I1347-02DUP	1	Solid	20.01	20	8.26	01/20/2017	10:51
8	I1347-04	1	Solid	20.04	20	7.79	01/20/2017	10:52
9	I1347-06	1	Solid	20.02	20	8.85	01/20/2017	10:53
10	I1347-08	1	Solid	20.02	100	9.92	01/20/2017	10:55
11	I1347-10	1	Solid	20.03	100	9.76	01/20/2017	10:56
12	I1347-12	1	Solid	20.01	100	9.90	01/20/2017	10:57
13	I1347-14	1	Solid	20.04	100	7.78	01/20/2017	10:58
14	I1347-16	1	Solid	20.05	100	7.71	01/20/2017	11:00
15	I1347-18	1	Solid	20.03	100	7.66	01/20/2017	11:01
16	CCV2	1	Water			12.02	01/20/2017	11:02
17	I1347-20	1	Solid	20.02	100	8.91	01/20/2017	11:03
18	I1351-01	1	Solid	20.02	100	9.69	01/20/2017	11:04
19	I1351-02	1	Solid	20.04	100	7.51	01/20/2017	11:05
20	CCV3	1	Water			12.02	01/20/2017	11:06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I1347

WorkList ID : 94671

Date : 1/20/2017 10:30:56 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/24/2017	Solid	I1347-02	Corrosivity	Cool 4 deg C	PSEG01	D52	BLACK-FILL1-0-5FT-COMP	01/19/2017	9045C
01/24/2017	Solid	I1347-04	Corrosivity	Cool 4 deg C	PSEG01	D52	BLACK-FILL2-0-5FT-COMP	01/19/2017	9045C
01/24/2017	Solid	I1347-06	Corrosivity	Cool 4 deg C	PSEG01	D52	S1-0-3FT-COMP	01/19/2017	9045C
01/24/2017	Solid	I1347-08	Corrosivity	Cool 4 deg C	PSEG01	D52	S2-0-3FT-COMP	01/19/2017	9045C
01/24/2017	Solid	I1347-10	Corrosivity	Cool 4 deg C	PSEG01	D52	S3-0-3FT-COMP	01/19/2017	9045C
01/24/2017	Solid	I1347-12	Corrosivity	Cool 4 deg C	PSEG01	D52	S4-0-3FT-COMP	01/19/2017	9045C
01/24/2017	Solid	I1347-14	Corrosivity	Cool 4 deg C	PSEG01	D52	T1-0-2FT-COMP	01/19/2017	9045C
01/24/2017	Solid	I1347-16	Corrosivity	Cool 4 deg C	PSEG01	D52	T2-0-2FT-COMP	01/19/2017	9045C
01/24/2017	Solid	I1347-18	Corrosivity	Cool 4 deg C	PSEG01	D52	S5-0-3FT-COMP	01/19/2017	9045C
01/24/2017	Solid	I1347-20	Corrosivity	Cool 4 deg C	PSEG01	D52	S6-0-3FT-COMP	01/19/2017	9045C
	Solid	I1351-01	Corrosivity	Cool 4 deg C	PARS15	D21	CAS-WCS-011917(0-15)	01/19/2017	9045C
	Solid	I1351-02	Corrosivity	Cool 4 deg C	PARS15	D21	CAS-WCS-011917(15-70)	01/19/2017	9045C

Date/Time 01-20-17 10:30 AM
 Received by: JH
 Relinquished by: Ca

Date/Time 01-20-17 weeph
 Received by: CR
 Relinquished by: JD