

7470A


LB8T452

INSTRUMENT ID: LV4

Linear

 μ Abs.:

38316



A= 0.0000e+000

B= 2.6289e-004

C= 6.1332e-003

Rho= 0.9998950

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	-6	0.000	-6				
0.2	0.200	0.220	0.020	815	0.0 %	815				
2.5	2.500	2.520	0.020	9562	0.0 %	9562				
5.0	5.000	4.973	-0.027	18893	0.0 %	18893				
7.5	7.500	7.403	-0.097	28138	0.0 %	28138				
10.0	10.000	10.079	0.079	38316	0.0 %	38316				

LB85452 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Jan 2017 10:21:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-6	-	0.00007470A	PPB		23 Jan 2017 10:51:22	S	Std
0.2 - 1	SO.2	815	-	0.20007470A	PPB		23 Jan 2017 10:53:31	S	Std
2.5 - 1	SO.5	9562	-	2.50007470A	PPB		23 Jan 2017 10:55:39	S	Std
5.0 - 1	SO	18893	-	5.00007470A	PPB		23 Jan 2017 10:57:47	S	Std
7.5 - 1	SO.5	28138	-	7.50007470A	PPB		23 Jan 2017 10:59:55	S	Std
10.0 - 1	SO	38316	-	10.00007470A	PPB		23 Jan 2017 11:02:02	S	Std
ICV06 - 1	ICV06	15409	4.0570	-7470A	PPB		23 Jan 2017 11:06:41	U	SMPL
ICB06 - 1	ICB06	-35	-0.0031	-7470A	PPB		23 Jan 2017 11:08:47	U	SMPL
CCV06 - 1	CCV06	18626	4.9027	-7470A	PPB		23 Jan 2017 11:10:53	U	SMPL
CCB06 - 1	CCB06	-29	-0.0015	-7470A	PPB		23 Jan 2017 11:13:00	U	SMPL
CRA - 1	CRA	798	0.2159	-7470A	PPB		23 Jan 2017 11:15:07	U	SMPL
HighStd - 1	HighStd	38036	10.0054	-7470A	PPB		23 Jan 2017 11:17:13	U	SMPL
ChkStd - 1	ChkStd	26456	6.9611	-7470A	PPB		23 Jan 2017 11:19:20	U	SMPL
PB96228BL - 1	PBW	-54	-0.0081	-7470A	PPB		23 Jan 2017 11:21:27	U	SMPL
PB96228BS - 1	LCSW	14381	3.7867	-7470A	PPB		23 Jan 2017 11:23:35	U	SMPL
PB96228TB - 1	PB96156TB	-31	-0.0020	-7470A	PPB		23 Jan 2017 11:25:42	U	SMPL
I1304-03 - 1	ROLL-OFF-15-41-COMP	-3	0.0053	-7470A	PPB		23 Jan 2017 11:27:51	U	SMPL
I1304-03DUP - 1	ROLL-OFF-15-41-COMPDUP	-1	0.0059	-7470A	PPB		23 Jan 2017 11:29:59	U	SMPL
I1304-03MS - 1	ROLL-OFF-15-41-COMPMS	13954	3.6745	-7470A	PPB		23 Jan 2017 11:32:08	U	SMPL
I1304-03MSD - 1	ROLL-OFF-15-41-COMPMSD	14004	3.6876	-7470A	PPB		23 Jan 2017 11:34:14	U	SMPL
CCV07 - 1	CCV07	18809	4.9508	-7470A	PPB		23 Jan 2017 11:36:21	U	SMPL
CCB07 - 1	CCB07	-40	-0.0044	-7470A	PPB		23 Jan 2017 11:38:28	U	SMPL
I1304-06 - 1	1634-NOTTINGHAM-COMP	3	0.0069	-7470A	PPB		23 Jan 2017 11:40:34	U	SMPL
I1304-09 - 1	ROLL-OFF-15939-COMP	12	0.0093	-7470A	PPB		23 Jan 2017 11:42:41	U	SMPL
I1308-04 - 1	MH-2-0-12FT-COMP	8	0.0082	-7470A	PPB		23 Jan 2017 11:44:48	U	SMPL
I1347-02 - 1	BLACK-FILL1-0-5FT-COMP	7	0.0080	-7470A	PPB		23 Jan 2017 11:46:55	U	SMPL
I1347-04 - 1	BLACK-FILL2-0-5FT-COMP	133	0.0411	-7470A	PPB		23 Jan 2017 11:49:03	U	SMPL
I1347-06 - 1	S1-0-3FT-COMP	20	0.0114	-7470A	PPB		23 Jan 2017 11:51:10	U	SMPL
I1347-08 - 1	S2-0-3FT-COMP	18	0.0109	-7470A	PPB		23 Jan 2017 11:53:18	U	SMPL
I1347-10 - 1	S3-0-3FT-COMP	38	0.0161	-7470A	PPB		23 Jan 2017 11:55:26	U	SMPL
I1347-12 - 1	S4-0-3FT-COMP	96	0.0314	-7470A	PPB		23 Jan 2017 11:57:35	U	SMPL
I1347-14 - 1	T1-0-2FT-COMP	19	0.0111	-7470A	PPB		23 Jan 2017 11:59:42	U	SMPL
CCV08 - 1	CCV08	18620	4.9011	-7470A	PPB		23 Jan 2017 12:01:49	U	SMPL
CCB08 - 1	CCB08	-31	-0.0020	-7470A	PPB		23 Jan 2017 12:03:56	U	SMPL
I1347-16 - 1	T2-0-2FT-COMP	2	0.0067	-7470A	PPB		23 Jan 2017 12:06:03	U	SMPL
I1347-18 - 1	S5-0-3FT-COMP	0	0.0061	-7470A	PPB		23 Jan 2017 12:08:10	U	SMPL
I1347-20 - 1	S6-0-3FT-COMP	-1	0.0059	-7470A	PPB		23 Jan 2017 12:10:17	U	SMPL
PB96230BL - 1	PBW	2	0.0067	-7470A	PPB		23 Jan 2017 12:12:24	U	SMPL
PB96230BS - 1	LCSW	14661	3.8603	-7470A	PPB		23 Jan 2017 12:14:32	U	SMPL
I1312-09 - 1	CAS-GW-02	423	0.1173	-7470A	PPB		23 Jan 2017 12:16:39	U	SMPL
I1312-10 - 1	CAS-WC011817-LIQUID	100	0.0324	-7470A	PPB		23 Jan 2017 12:18:47	U	SMPL
I1312-10DUP - 1	CAS-WC011817-LIQUIDDUP	62	0.0224	-7470A	PPB		23 Jan 2017 12:20:55	U	SMPL
I1312-10MS - 1	CAS-WC011817-LIQUIDMS	16459	4.3330	-7470A	PPB		23 Jan 2017 12:23:04	U	SMPL
I1312-10MSD - 1	CAS-WC011817-LIQUIDMSD	15714	4.1372	-7470A	PPB		23 Jan 2017 12:25:10	U	SMPL
CCV09 - 1	CCV09	18871	4.9671	-7470A	PPB		23 Jan 2017 12:27:18	U	SMPL
CCB09 - 1	CCB09	-32	-0.0023	-7470A	PPB		23 Jan 2017 12:29:25	U	SMPL
I1351-04 - 1	CAS-FB011917	-7	0.0043	-7470A	PPB		23 Jan 2017 12:31:32	U	SMPL
I1304-03LX5 - 1		-17	0.0017	-7470A	PPB		23 Jan 2017 12:33:39	U	SMPL
I1304-03A - 1		15281	4.0233	-7470A	PPB		23 Jan 2017 12:35:46	U	SMPL
I1312-10LX5 - 1		-3	0.0053	-7470A	PPB		23 Jan 2017 12:37:53	U	SMPL
I1312-10A - 1		15091	3.9734	-7470A	PPB		23 Jan 2017 12:40:01	U	SMPL
CCV10 - 1	CCV10	19017	5.0055	-7470A	PPB		23 Jan 2017 12:42:09	U	SMPL
CCB10 - 1	CCB10	-42	-0.0049	-7470A	PPB		23 Jan 2017 12:44:17	U	SMPL