



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85456

Temp.: 22.3 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/23/2017	11:00
2	CAL2	1	Water			7.01	01/23/2017	11:01
3	CAL3	1	Water			10.04	01/23/2017	11:02
4	ICV	1	Water			6.99	01/23/2017	11:03
5	CCV1	1	Water			2.02	01/23/2017	11:04
6	I1379-06	1	Solid	20.02	20	8.73	01/23/2017	11:05
7	I1379-06DUP	1	Solid	20.04	20	8.74	01/23/2017	11:06
8	CCV2	1	Water			12.01	01/23/2017	11:07

IB 85456

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I1379 WorkList ID : 94720 Date : 1/23/2017 10:42:42 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/03/2017	Solid	I1379-06	Corrosivity	Cool 4 deg C	DVIR01	D22	WC-COMPOSITE	01/20/2017	9045C

Date/Time 01-23-17 10:40 AM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 01-23-17 12:10 PM
Received by: [Signature]
Relinquished by: 51