



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85461

Temp.: 22.3 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/23/2017	11:00
2	CAL2	1	Water			7.01	01/23/2017	11:01
3	CAL3	1	Water			10.04	01/23/2017	11:02
4	ICV	1	Water			7.00	01/23/2017	11:03
5	CCV1	1	Water			2.02	01/23/2017	11:04
6	I1272-06	1	Solid	20.02	20	3.10	01/23/2017	11:05
7	CCV2	1	Water			12.02	01/23/2017	11:06