

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/24/17

Analyst : AK

Data File : LB85465.csv

Approved By : *nb*

Approved Date : *1/24/17*

Worksheet # :


m9012 A-B Total Amenable & Reactive Cyanide -16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	1/23/17						
Cyanide		PASS	1.534		0.002						mg/L	
5.0PPBCN	5.0PPBCN				W	1/23/17						
Cyanide		PASS	6.162		0.006						mg/L	
10PPBCN	10PPBCN				W	1/23/17						
Cyanide		PASS	10.576		0.011						mg/L	
50PPBCN	50PPBCN				W	1/23/17						
Cyanide		PASS	52.431		0.052						mg/L	
100PPBCN	100PPBCN				W	1/23/17						
Cyanide		PASS	96.384		0.096						mg/L	
250PPBCN	250PPBCN				W	1/23/17						
Cyanide		PASS	244.911		0.245						mg/L	
500PPBCN	500PPBCN				W	1/23/17						
Cyanide		PASS	503.001		0.503						mg/L	
ICV1	ICV1				W	1/23/17						
Cyanide		PASS	93.209		0.09	90.0	90	110			mg/L	
ICB1	ICB1				W	1/23/17						
Cyanide		PASS	2.010		0.002			+/-0.0030			mg/L	
CCV1	CCV1				W	1/23/17						
Cyanide		PASS	235.568		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	1/23/17						
Cyanide		PASS	1.969		0.002			+/-0.0030			mg/L	
PB96244BLS	PB96244BLS				S	1/23/17						
Cyanide		PASS	1.802		0.090			+/-0.0330			mg/Kg	
PB96244BSS	PB96244BSS				S	1/23/17						
Cyanide		PASS	94.055		4.70	94.0	85.00	115.00			mg/Kg	
LOWPB96244	LOWPB96244				W	1/23/17						
Cyanide		PASS	11.173		0.011						mg/L	
HIGHPB96244	HIGHPB96244				W	1/23/17						
Cyanide		PASS	492.743		0.493						mg/L	
I1326-01	FK-GF-02-001-A			1	S	1/23/17						
Cyanide		PASS	2.545		0.127						mg/Kg	
I1326-03	FK-GF-02-001-B			1	S	1/23/17						
Cyanide		PASS	2.034		0.101						mg/Kg	
I1326-05	FK-GF-02-001-C			1	S	1/23/17						
Cyanide		PASS	2.107		0.103						mg/Kg	
I1329-01	BLACK-FILL1-0-5FT-COMP			1	S	1/23/17						
Cyanide		PASS	9.569		0.701						mg/Kg	
I1329-01D	BLACK-FILL1-0-5FT-COMP			1	S	1/23/17						
Cyanide		PASS	10.290		0.753				7.2	20	mg/Kg	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/24/17
 Analyst : AK
 Data File : LB85465.csv

Approved By : 
 Approved Date :
 Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep Method	Analysis Method				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide											
I1329-01S	BLACK-FILL1-0-5FT-COMPS	1	S	1/23/17							
Cyanide	PASS	48.489	3.590	98.0	75	125				mg/Kg	
CCV2	CCV2		W	1/23/17							
Cyanide	PASS	237.914	0.24	96.0	90	110				mg/L	
CCB2	CCB2		W	1/23/17							
Cyanide	PASS	1.839	0.002				+/-0.0030			mg/L	
I1329-01SD	BLACK-FILL1-0-5FT-COMPS	1	S	1/23/17							
Cyanide	PASS	48.841	3.610	98.0	75	125	0.6	20		mg/Kg	
I1329-03	BLACK-FILL2-0-5FT-COMP	1	S	1/23/17							
Cyanide	PASS	13.929	0.899							mg/Kg	
I1329-05	S1-0-3FT-COMP	1	S	1/23/17							
Cyanide	PASS	6.016	0.357							mg/Kg	
I1329-07	S2-0-3FT-COMP	1	S	1/23/17							
Cyanide	PASS	3.955	0.234							mg/Kg	
I1329-09	S3-0-3FT-COMP	1	S	1/23/17							
Cyanide	PASS	6.780	0.391							mg/Kg	
I1329-11	S4-0-3FT-COMP	1	S	1/23/17							
Cyanide	PASS	3.622	0.214							mg/Kg	
I1329-13	T1-0-2FT-COMP	1	S	1/23/17							
Cyanide	PASS	-0.354	-0.038							mg/Kg	
I1329-15	T2-0-2FT-COMP	1	S	1/23/17							
Cyanide	PASS	1.330	0.113							mg/Kg	
I1329-17	S5-0-3FT-COMP	1	S	1/23/17							
Cyanide	PASS	3.024	0.172							mg/Kg	
I1329-19	S6-0-3FT-COMP	1	S	1/23/17							
Cyanide	PASS	2.070	0.117							mg/Kg	
CCV3	CCV3		W	1/23/17							
Cyanide	PASS	241.477	0.24	96.0	90	110				mg/L	
CCB3	CCB3		W	1/23/17							
Cyanide	PASS	1.801	0.002				+/-0.0030			mg/L	