



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85470

Temp.: 19.4 °C

BalanceID:

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	01/24/2017	12:10
2	CAL2	1	Water			7.02	01/24/2017	12:11
3	CAL3	1	Water			10.03	01/24/2017	12:12
4	ICV	1	Water			7.01	01/24/2017	12:13
5	CCV1	1	Water			2.02	01/24/2017	12:15
6	I1422-01	1	Water			5.03	01/24/2017	12:16
7	I1422-01DUP	1	Water			5.02	01/24/2017	12:17
8	I1422-02	1	Water			4.83	01/24/2017	12:18
9	CCV2	1	Water			12.02	01/24/2017	12:19

WORKLIST(Hardcopy Internal Chain)

LB 85470

WorkList Name : PH 11422 WorkList ID : 94792 Date : 1/24/2017 12:01:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1422-01	pH	Cool 4 deg C	STER07	E11	CELL-A	01/19/2017	SM 4500-PH B
	Water	I1422-02	pH	Cool 4 deg C	STER07	E11	CELLS-B-C	01/19/2017	SM 4500-PH B

Date/Time 01-24-17 12:01 PM
 Received by: JP
 Relinquished by: JP

Date/Time 01-24-17 11:30 PM
 Received by: JP
 Relinquished by: JP