


7471A

L985472
INSTRUMENT ID: CV1

Linear

 μ Abs.:

4551E



A= 0.0000e+000

B= 2.2105e-004

C= 8.1183e-004

Rho= 0.9998731

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.000	-0.000	-5	0.000	-5				
0.2	0.200	0.199	-0.001	896	0.0 %	896				
2.5	2.500	2.515	0.015	11373	0.0 %	11373				
5.0	5.000	5.045	0.045	22819	0.0 %	22819				
7.5	7.500	7.379	-0.121	33378	0.0 %	33378				
10.0	10.000	10.062	0.062	45516	0.0 %	45516				

LB85472 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 24 Jan 2017 12:45:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-5	-	0.0000	7471A	PPB	24 Jan 2017 12:52:48	S	Std
0.2 - 1	502	896	-	0.2000	7471A	PPB	24 Jan 2017 12:55:06	S	Std
2.5 - 1	525	11373	-	2.5000	7471A	PPB	24 Jan 2017 12:57:23	S	Std
5.0 - 1	550	22819	-	5.0000	7471A	PPB	24 Jan 2017 13:02:21	S	Std
7.5 - 1	575	33378	-	7.5000	7471A	PPB	24 Jan 2017 13:04:38	S	Std
10.0 - 1	510	45516	-	10.0000	7471A	PPB	24 Jan 2017 13:09:35	S	Std
ICV07 - 1	ICV07	16685	3.6891	-7471A	PPB	24 Jan 2017 13:12:17	U	SMPL	
ICB07 - 1	ICB07	-33	-0.0065	-7471A	PPB	24 Jan 2017 13:14:33	U	SMPL	
CCV11 - 1	CCV11	22450	4.9635	-7471A	PPB	24 Jan 2017 13:16:49	U	SMPL	
CCB11 - 1	CCB11	-43	-0.0087	-7471A	PPB	24 Jan 2017 13:19:04	U	SMPL	
CRA - 1	CRA	883	0.1960	-7471A	PPB	24 Jan 2017 13:21:20	U	SMPL	
HighStd - 1	HighStd	45723	10.1081	-7471A	PPB	24 Jan 2017 13:23:36	U	SMPL	
ChkStd - 1	ChkStd	30691	6.7852	-7471A	PPB	24 Jan 2017 13:25:51	U	SMPL	
PB96256BL - 1	PBS	-52	-0.0107	-7471A	PPB	24 Jan 2017 13:28:10	U	SMPL	
PB96256BS - 1	LCSS	15088	3.3361	-7471A	PPB	24 Jan 2017 13:30:26	U	SMPL	
I1329-01 - 1	BLACK-FILL1-0-5FT-COMP	35107	7.7614	-7471A	PPB	24 Jan 2017 13:32:43	U	SMPL	
I1329-03 - 1	BLACK-FILL2-0-5FT-COMP	46840	10.3550	-7471A	PPB	24 Jan 2017 13:35:00	U	SMPL	
I1329-05 - 1	S1-0-3FT-COMP	25502	5.6381	-7471A	PPB	24 Jan 2017 13:37:18	U	SMPL	
I1329-07 - 1	S2-0-3FT-COMP	27588	6.0993	-7471A	PPB	24 Jan 2017 13:39:35	U	SMPL	
I1329-09 - 1	S3-0-3FT-COMP	50421	11.1466	-7471A	PPB	24 Jan 2017 13:41:52	U	SMPL	
CCV12 - 1	CCV12	21893	4.8404	-7471A	PPB	24 Jan 2017 13:44:08	U	SMPL	
CCB12 - 1	CCB12	-45	-0.0091	-7471A	PPB	24 Jan 2017 13:46:24	U	SMPL	
I1329-11 - 1	S4-0-3FT-COMP	28765	6.3594	-7471A	PPB	24 Jan 2017 13:48:39	U	SMPL	
I1329-13 - 1	T1-0-2FT-COMP	3506	0.7758	-7471A	PPB	24 Jan 2017 13:50:56	U	SMPL	
I1329-15 - 1	T2-0-2FT-COMP	9572	2.1167	-7471A	PPB	24 Jan 2017 13:53:12	U	SMPL	
I1329-17 - 1	S5-0-3FT-COMP	14548	3.2167	-7471A	PPB	24 Jan 2017 13:55:28	U	SMPL	
I1329-19 - 1	S6-0-3FT-COMP	14374	3.1782	-7471A	PPB	24 Jan 2017 13:57:44	U	SMPL	
I1349-01 - 1	WASTE	159	0.0360	-7471A	PPB	24 Jan 2017 14:00:01	U	SMPL	
I1349-01DUP - 1	WASTEDUP	176	0.0397	-7471A	PPB	24 Jan 2017 14:02:18	U	SMPL	
I1349-01MS - 1	WASTEMS	16217	3.5857	-7471A	PPB	24 Jan 2017 14:04:35	U	SMPL	
I1349-01MSD - 1	WASTEMSD	16052	3.5492	-7471A	PPB	24 Jan 2017 14:06:53	U	SMPL	
I1351-03 - 1	CAS-SB-01-68-70	121	0.0276	-7471A	PPB	24 Jan 2017 14:09:09	U	SMPL	
CCV13 - 1	CCV13	22158	4.8989	-7471A	PPB	24 Jan 2017 14:11:25	U	SMPL	
CCB13 - 1	CCB13	-27	-0.0052	-7471A	PPB	24 Jan 2017 14:13:41	U	SMPL	
I1363-01 - 1	BLDG-3-FLOOR-COMP	178	0.0402	-7471A	PPB	24 Jan 2017 14:15:57	U	SMPL	
I1363-04 - 1	BLDG-3-BLOCK-WALL-COMP	182	0.0410	-7471A	PPB	24 Jan 2017 14:18:13	U	SMPL	
I1363-07 - 1	BLDG-1-FLOOR-COMP	176	0.0397	-7471A	PPB	24 Jan 2017 14:20:29	U	SMPL	
PB96257BL - 1	PBS	0	0.0008	-7471A	PPB	24 Jan 2017 14:22:46	U	SMPL	
PB96257BS - 1	LCSS	15242	3.3701	-7471A	PPB	24 Jan 2017 14:25:03	U	SMPL	
I1378-01 - 1	RVE-1	5628	1.2449	-7471A	PPB	24 Jan 2017 14:27:19	U	SMPL	
I1378-02 - 1	RVE-2	3643	0.8061	-7471A	PPB	24 Jan 2017 14:29:36	U	SMPL	
I1378-03 - 1	RVE-3	1048	0.2325	-7471A	PPB	24 Jan 2017 14:31:53	U	SMPL	
I1378-04 - 1	RVE-4	1344	0.2979	-7471A	PPB	24 Jan 2017 14:34:11	U	SMPL	
I1378-05 - 1	RVE-6	1576	0.3492	-7471A	PPB	24 Jan 2017 14:36:27	U	SMPL	
CCV14 - 1	CCV14	22691	5.0168	-7471A	PPB	24 Jan 2017 14:38:43	U	SMPL	
CCB14 - 1	CCB14	-39	-0.0078	-7471A	PPB	24 Jan 2017 14:41:00	U	SMPL	
I1378-06 - 1	RVE-7	862	0.1914	-7471A	PPB	24 Jan 2017 14:43:16	U	SMPL	
I1378-07 - 1	RVE-8	571	0.1270	-7471A	PPB	24 Jan 2017 14:45:33	U	SMPL	
I1378-08 - 1	RVE-9	4609	1.0197	-7471A	PPB	24 Jan 2017 14:47:49	U	SMPL	
I1378-09 - 1	RVE-10	1471	0.3260	-7471A	PPB	24 Jan 2017 14:50:05	U	SMPL	
I1379-05 - 1	WC-COMPOSITE	3693	0.8172	-7471A	PPB	24 Jan 2017 14:52:21	U	SMPL	
I1382-01 - 1	FK-GF-01-001-A	485	0.1080	-7471A	PPB	24 Jan 2017 14:54:38	U	SMPL	
I1382-03 - 1	FK-GF-01-001-B	1715	0.3799	-7471A	PPB	24 Jan 2017 14:56:55	U	SMPL	
I1382-05 - 1	FK-GF-01-001-C	1742	0.3859	-7471A	PPB	24 Jan 2017 14:59:12	U	SMPL	
I1383-01 - 1	FK-SF-01-001-A	1016	0.2254	-7471A	PPB	24 Jan 2017 15:01:30	U	SMPL	
I1383-03 - 1	FK-SF-01-001-B	1163	0.2579	-7471A	PPB	24 Jan 2017 15:03:46	U	SMPL	
CCV15 - 1	CCV15	22443	4.9619	-7471A	PPB	24 Jan 2017 15:06:03	U	SMPL	
CCB15 - 1	CCB15	-35	-0.0069	-7471A	PPB	24 Jan 2017 15:08:19	U	SMPL	
I1383-05 - 1	FK-SF-01-001-C	1022	0.2267	-7471A	PPB	24 Jan 2017 15:10:35	U	SMPL	
I1383-05DUP - 1	FK-SF-01-001-CDUP	1151	0.2552	-7471A	PPB	24 Jan 2017 15:12:51	U	SMPL	
I1383-05MS - 1	FK-SF-01-001-CMS	16563	3.6621	-7471A	PPB	24 Jan 2017 15:15:07	U	SMPL	
I1383-05MSD - 1	FK-SF-01-001-CMSD	16653	3.6820	-7471A	PPB	24 Jan 2017 15:17:24	U	SMPL	
I1349-01LX5 - 1		5	0.0019	-7471A	PPB	24 Jan 2017 15:19:40	U	SMPL	
I1349-01A - 1		18126	4.0076	-7471A	PPB	24 Jan 2017 15:21:56	U	SMPL	
I1383-05LX5 - 1		159	0.0360	-7471A	PPB	24 Jan 2017 15:24:14	U	SMPL	
I1383-05A - 1		16473	3.6422	-7471A	PPB	24 Jan 2017 15:26:31	U	SMPL	
I1329-03DLX5 - 1		8898	1.9678	-7471A	PPB	24 Jan 2017 15:28:49	U	SMPL	
I1329-09DLX5 - 1		9478	2.0960	-7471A	PPB	24 Jan 2017 15:31:06	U	SMPL	
CCV16 - 1	CCV16	22210	4.9104	-7471A	PPB	24 Jan 2017 15:33:22	U	SMPL	
CCB16 - 1	CCB16	-32	-0.0063	-7471A	PPB	24 Jan 2017 15:35:39	U	SMPL	