

Analysis Method: SM4500-CL C

ANALYST: AMANDEEP

Parameter: Chloride

SUPERVISOR REVIEW BY: apatel

Run Number: LB85489

Constant: 35450

Normality1: 0.0141

Reagent/Standard	Lot/Log #
Chloride LCS std, 100PPM	WP51235
Indicator-acidifier reagent - Chloride	WP51766
MERCURIC NITRATE, 0.0141 N 4L	W2051

Seq	Lab ID	True Value (mg/L)	Dilution	Sample Volume (ml)	Titrant Initial (ml)	Titrant Final (ml)	Titrant Diff. (ml)	Result (mg/L)	Anal Date	Anal Time
1	LB85489BL		1	50	0.00	0.04	0.04	0.00	01/24/2017	10:56
2	LB85489BS	100	1	10	0.04	2.04	2.00	97.97	01/24/2017	11:04
3	I1272-08		1	5	2.04	3.58	1.54	149.95	01/24/2017	11:10

Titrant Diff = TitrantFinal - TitrantInitial

Result = (Titrant Diff - Row Blank Value * Normality)) * Constant /