



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85491

Temp.: 21.4 °C

BalanceID:

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	01/25/2017	12:40
2	CAL2	1	Water			7.01	01/25/2017	12:41
3	CAL3	1	Water			10.02	01/25/2017	12:42
4	ICV	1	Water			7.01	01/25/2017	12:43
5	CCV1	1	Water			2.01	01/25/2017	12:44
6	I1448-09	1	Solid	20.02	20	5.16	01/25/2017	12:45
7	I1448-11	1	Solid	20.03	20	8.73	01/25/2017	12:46
8	I1448-11DUP	1	Solid	20.01	20	8.74	01/25/2017	12:47
9	CCV2	1	Water			12.02	01/25/2017	12:48

21385491

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I1448 WorkList ID : 94834 Date : 1/25/2017 12:30:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1448-09	Corrosivity	Cool 4 deg C	PARS15		CHA-WCS-012317-1(0-70)	01/23/2017	9045C
	Solid	I1448-11	Corrosivity	Cool 4 deg C	PARS15		CHA-WCS-012417-1(0-4)	01/24/2017	9045C

Date/Time 01-25-17 12:30 PM
Received by: SP
Relinquished by:

Date/Time 01-25-17 2:00 PM
Received by: SP
Relinquished by: SP