



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB85511

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	01/26/2017	11:30
2	I1312-13		1	>100.0	>212.0	01/26/2017	11:50
3	I1312-13DUP		1	>100.0	>212.0	01/26/2017	12:15

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 85571

WorkList Name : FP12617

WorkList ID : 94878

Date : 1/26/2017 10:59:33 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1312-13	Flash Point	Cool 4 deg C	PARS15	D11	CAS-WC011817-LIQUID	01/18/2017	1010A

Date/Time 1/26/17 11:00 AM
 Received by: KS
 Relinquished by: CS

Date/Time 1/26/17 12:20 PM
 Received by: CS
 Relinquished by: KS