

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/27/17
 Analyst :
 Data File : LB85529.csv

Approved By :
 Approved Date :
 Worksheet # :

M300-0 - Investigative Analysis - IL

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Bromide												
STD1	STD1				W	1/16/17						
Bromide		PASS	0.000		0.000						mg/L	
STD2	STD2				W	1/16/17						
Bromide		PASS	0.913		0.913						mg/L	
STD3	STD3				W	1/16/17						
Bromide		PASS	2.270		2.27						mg/L	
STD4	STD4				W	1/16/17						
Bromide		PASS	4.082		4.08						mg/L	
STD5	STD5				W	1/16/17						
Bromide		PASS	9.756		9.76						mg/L	
STD6	STD6				W	1/16/17						
Bromide		PASS	19.527		19.5						mg/L	
STD7	STD7				W	1/16/17						
Bromide		PASS	24.510		24.5						mg/L	
STD8	STD8				W	1/16/17						
Bromide		PASS	50.442		50.4						mg/L	
ICV1	ICV1				W	1/16/17						
Bromide		PASS	25.395		25.40	102.0	90	110			mg/L	
ICB1	ICB1				W	1/16/17						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	
CCV1	CCV1				W	1/24/17						
Bromide		PASS	25.734		25.70	103.0	90	110			mg/L	
CCB1	CCB1				W	1/24/17						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	
LB85529BLW	LB85529BLW				W	1/24/17						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	
LB85529BSW	LB85529BSW				W	1/24/17						
Bromide		PASS	25.914		25.90	104.0	90	110			mg/L	
I1422-01	CELL-A		1		W	1/24/17						
Bromide		PASS	0.000		0.000						mg/L	
I1422-02	CELLS-B-C		1		W	1/24/17						
Bromide		PASS	0.000		0.000						mg/L	
I1422-02D	CELLS-B-CD		1		W	1/24/17						
Bromide		PASS	0.000		0.00				0	20	mg/L	
I1422-02S	CELLS-B-CS		1		W	1/24/17						
Bromide		PASS	24.186		24.20	97.0	80	120			mg/L	
CCV2	CCV2				W	1/24/17						
Bromide		PASS	24.946		24.90	100.0	90	110			mg/L	
CCB2	CCB2				W	1/24/17						
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/27/17
 Analyst :
 Data File : LB85529.csv

Approved By :
 Approved Date : 1/27/17
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis		Line 1
Parameter			PPB	Final Conc		%Rec	Method	Method		Line 2
Chloride										
STD1	STD1				W	1/16/17				
Chloride		PASS	0.000	0.000					mg/L	
STD2	STD2				W	1/16/17				
Chloride		PASS	0.357	0.357					mg/L	
STD3	STD3				W	1/16/17				
Chloride		PASS	0.765	0.765					mg/L	
STD4	STD4				W	1/16/17				
Chloride		PASS	1.259	1.26					mg/L	
STD5	STD5				W	1/16/17				
Chloride		PASS	2.867	2.87					mg/L	
STD6	STD6				W	1/16/17				
Chloride		PASS	5.722	5.72					mg/L	
STD7	STD7				W	1/16/17				
Chloride		PASS	7.248	7.25					mg/L	
STD8	STD8				W	1/16/17				
Chloride		PASS	15.231	15.2					mg/L	
ICV1	ICV1				W	1/16/17				
Chloride		PASS	7.295	7.30		97.0	90	110	mg/L	
ICB1	ICB1				W	1/16/17				
Chloride		PASS	0.000	0.000				+/-0.0750	mg/L	
CCV1	CCV1				W	1/24/17				
Chloride		PASS	7.441	7.44		99.0	90	110	mg/L	
CCB1	CCB1				W	1/24/17				
Chloride		PASS	0.000	0.000				+/-0.0750	mg/L	
LB85529BLW	LB85529BLW				W	1/24/17				
Chloride		PASS	0.000	0.000				+/-0.0750	mg/L	
LB85529BSW	LB85529BSW				W	1/24/17				
Chloride		PASS	7.592	7.59		101.0	90	110	mg/L	
I1422-01	CELL-A			1	W	1/24/17				
Chloride		PASS	0.672	0.672					mg/L	
I1422-02	CELLS-B-C			1	W	1/24/17				
Chloride		PASS	0.000	0.000					mg/L	
I1422-02D	CELLS-B-CD			1	W	1/24/17				
Chloride		PASS	0.000	0.00				0	20	mg/L
I1422-02S	CELLS-B-CS			1	W	1/24/17				
Chloride		PASS	7.150	7.15		95.0	80	120	mg/L	
CCV2	CCV2				W	1/24/17				
Chloride		PASS	7.324	7.32		98.0	90	110	mg/L	
CCB2	CCB2				W	1/24/17				
Chloride		PASS	0.000	0.000				+/-0.0750	mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/27/17
 Analyst : 
 Data File : LB85529.csv

Approved By : 
 Approved Date : 1/27/17
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method			Line 2
Fluoride											
STD1	STD1				W	1/16/17					
Fluoride		PASS	0.000		0.000					mg/L	
STD2	STD2				W	1/16/17					
Fluoride		PASS	0.255		0.255					mg/L	
STD3	STD3				W	1/16/17					
Fluoride		PASS	0.470		0.470					mg/L	
STD4	STD4				W	1/16/17					
Fluoride		PASS	0.845		0.845					mg/L	
STD5	STD5				W	1/16/17					
Fluoride		PASS	1.854		1.85					mg/L	
STD6	STD6				W	1/16/17					
Fluoride		PASS	3.888		3.89					mg/L	
STD7	STD7				W	1/16/17					
Fluoride		PASS	4.862		4.86					mg/L	
STD8	STD8				W	1/16/17					
Fluoride		PASS	10.125		10.1					mg/L	
ICV1	ICV1				W	1/16/17					
Fluoride		PASS	4.693		4.69	94.0	90	110		mg/L	
ICB1	ICB1				W	1/16/17					
Fluoride		PASS	0.000		0.000			+/-0.0430		mg/L	
CCV1	CCV1				W	1/24/17					
Fluoride		PASS	5.166		5.17	103.0	90	110		mg/L	
CCB1	CCB1				W	1/24/17					
Fluoride		PASS	0.000		0.000			+/-0.0430		mg/L	
LB85529BLW	LB85529BLW				W	1/24/17					
Fluoride		PASS	0.000		0.000			+/-0.0430		mg/L	
LB85529BSW	LB85529BSW				W	1/24/17					
Fluoride		PASS	5.256		5.26	105.0	90	110		mg/L	
I1422-01	CELL-A			1	W	1/24/17					
Fluoride		PASS	0.287		0.287					mg/L	
I1422-02	CELLS-B-C			1	W	1/24/17					
Fluoride		PASS	0.000		0.000					mg/L	
I1422-02D	CELLS-B-CD			1	W	1/24/17					
Fluoride		PASS	0.000		0.00			0	20	mg/L	
I1422-02S	CELLS-B-CS			1	W	1/24/17					
Fluoride		PASS	4.904		4.90	98.0	80	120		mg/L	
CCV2	CCV2				W	1/24/17					
Fluoride		PASS	5.067		5.07	101.0	90	110		mg/L	
CCB2	CCB2				W	1/24/17					
Fluoride		PASS	0.000		0.000			+/-0.0430		mg/L	