

Run Number: LB85532

DATE & TIME IN: 01/27/2017 10:00

OVEN TEMP: 105 °C

DATE & TIME OUT: 01/27/2017 14:10

OVEN TEMP: 104 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
1	LB85532BL	LB85532BL	1.3205	1.3205	100	1.3205	1.3205	1.3205	0	0
2	LB85532BS	LB85532BS	1.4685	1.4685	100	1.5229	1.5229	1.5229	0.0544	544
3	I1450-01	COMP-001	1.3976	1.3976	200	1.4101	1.4101	1.4101	0.0125	62.5
4	I1450-01DUP	COMP-001DUP	1.4114	1.4114	200	1.4245	1.4245	1.4245	0.0131	65.5
5	I1457-01	BAYPARK-122816	1.3999	1.3999	300	1.4084	1.4084	1.4084	0.0085	28.3
6	I1502-01	001M-WILLETS-PTBLVD (JAN)	1.4144	1.4144	100	1.4263	1.4263	1.4263	0.0119	119
7	I1502-02	002M-35THAVE (JAN)	1.3975	1.3975	100	1.4117	1.4117	1.4117	0.0142	142
8	I1508-01	SPDES-1	1.3952	1.3952	800	1.3954	1.3954	1.3954	0.0002	0.3
9	I1510-01	DRGFB-WQ-002-BG-A	1.4044	1.4044	300	1.4179	1.4179	1.4179	0.0135	45
10	I1510-02	DRGFB-WQ-002-BG-B	1.3993	1.3834	300	1.416	1.416	1.416	0.0326	108.7
11	I1510-03	DRGFB-WQ-002-BG-C	1.3834	1.3834	300	1.3961	1.3961	1.3961	0.0127	42.3
12	I1510-04	DRGFB-WQ-003-AD-A	1.414	1.414	300	1.4377	1.4377	1.4377	0.0237	79
13	I1510-05	DRGFB-WQ-005-BG-A	1.4095	1.4095	300	1.4158	1.4158	1.4158	0.0063	21

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

$$\text{Weight (g)} = C - B$$

$$\text{Result mg/L} = \frac{D}{A} * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS I1508

WorkList ID : 94909

Date : 1/27/2017 8:37:52 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1450-01	TSS	Cool 4 deg C	ARAM01	E12	COMP-001	01/25/2017	SM2540 D
	Water	I1457-01	TSS	Cool 4 deg C	BSRJ01	E12	BAYPARK-122816	01/24/2017	SM2540 D
	Water	I1502-01	TSS	Cool 4 deg C	TULL01	F11	001M-WILLETTS-PTBLVD(JAN)	01/25/2017	SM2540 D
	Water	I1502-02	TSS	Cool 4 deg C	TULL01	F11	002M-35THAVE(JAN)	01/25/2017	SM2540 D
	Water	I1508-01	TSS	Cool 4 deg C	SKAN06	F11	SPDES-1	01/26/2017	SM2540 D
	Water	I1510-01	TSS	Cool 4 deg C	JAYC01	F12	DRGFB-WQ-002-BG-A	01/25/2017	SM2540 D
	Water	I1510-02	TSS	Cool 4 deg C	JAYC01	F12	DRGFB-WQ-002-BG-B	01/25/2017	SM2540 D
	Water	I1510-03	TSS	Cool 4 deg C	JAYC01	F12	DRGFB-WQ-002-BG-C	01/25/2017	SM2540 D
	Water	I1510-04	TSS	Cool 4 deg C	JAYC01	F12	DRGFB-WQ-003-AD-A	01/25/2017	SM2540 D
	Water	I1510-05	TSS	Cool 4 deg C	JAYC01	F12	DRGFB-WQ-005-BG-A	01/26/2017	SM2540 D

Date/Time 01-27-16 8:37 AM
 Received by: JY
 Relinquished by: CS

Date/Time 01-27-16 12:10 PM
 Received by: CP
 Relinquished by: JP