



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB85534

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP52486
Residual chlorine std, 0PPM	WP52482
Residual chlorine std, 0.1PPM	WP52483
Residual chlorine LOQ, 0.2PPM	WP52490
Residual chlorine std, 0.2PPM	WP52484
Residual chlorine std, and CCV , 0.4PPM	WP52485
Residual chlorine std, 1.6PPM	WP52487
Residual chlorine ICV-LCS, 0.4PPM	WP52488
Residual chlorine LOD, 0.05PPM	WP52489
Total Chlorine Powder Pillows	W1960

Intercept: 0.0074

Slope: 1.0276

Regression: 0.999971

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	01/27/2017	08:45
2	CAL2	0.1	1	0.000	0.110	0.110	01/27/2017	08:50
3	CAL3	0.2	1	0.000	0.220	0.220	01/27/2017	08:55
4	CAL4	0.4	1	0.000	0.420	0.420	01/27/2017	09:00
5	CAL5	0.8	1	0.000	0.830	0.830	01/27/2017	09:05
6	CAL6	1.6	1	0.000	1.650	1.650	01/27/2017	09:10



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB85534

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.410	0.410	0.392	01/27/2017	09:15
2	ICB				1	0.000	0.000	0.000	-0.007	01/27/2017	09:20
3	CCV1			0.4	1	0.020	0.430	0.410	0.392	01/27/2017	09:25
4	CCB1				1	0.000	0.000	0.000	-0.007	01/27/2017	09:30
5	LB85534BL	50	50		1	0.000	0.000	0.000	-0.007	01/27/2017	09:35
6	LB85534BS	50	50	0.4	1	0.000	0.400	0.400	0.382	01/27/2017	09:40
7	I1092-03	50	50		1	0.000	0.050	0.050	0.041	01/27/2017	09:45
8	I1092-04	50	50		1	0.000	0.220	0.220	0.207	01/27/2017	09:50
9	CCV2			0.4	1	0.000	0.410	0.410	0.392	01/27/2017	09:55
10	CCB2				1	0.000	0.000	0.000	-0.007	01/27/2017	10:00

LB 85534

WORKLIST(Hardcopy Internal Chain)

WorkList Name : RCL011817

WorkList ID : 94539

Date : 1/18/2017 9:24:58 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1092-03	Residual Chlorine	Cool 4 deg C	CHEM02	Wet Chem L	Annual -LOD- Water-2017	01/09/2017	SM4500 Cl G
	Water	I1092-04	Residual Chlorine	Cool 4 deg C	CHEM02	Wet Chem L	Annual -LOQ- Water-2017	01/09/2017	SM4500 Cl G

Date/Time _____
Received by: _____
Relinquished by: _____

Date/Time _____
Received by: _____
Relinquished by: _____