



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85538

Temp.: 22.3 °C

BalanceID:

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.01 Control Limits[+/- 0.1].

True Value of CCV4 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/27/2017	11:20
2	CAL2	1	Water			7.01	01/27/2017	11:21
3	CAL3	1	Water			10.05	01/27/2017	11:22
4	ICV	1	Water			7.02	01/27/2017	11:23
5	CCV1	1	Water			2.02	01/27/2017	11:25
6	I1527-21	1	Solid	20.02	20	6.60	01/27/2017	11:26
7	I1527-21DUP	1	Solid	20.03	20	6.61	01/27/2017	11:27
8	I1527-22	1	Solid	20.05	20	7.06	01/27/2017	11:28
9	I1527-23	1	Solid	20.01	20	7.78	01/27/2017	11:29
10	I1527-24	1	Solid	20.02	20	7.73	01/27/2017	11:30
11	I1527-25	1	Solid	20.04	20	6.76	01/27/2017	11:31
12	I1527-26	1	Solid	20.01	20	7.36	01/27/2017	11:32
13	I1527-27	1	Solid	20.03	20	7.57	01/27/2017	11:33
14	I1527-28	1	Solid	20.04	20	7.83	01/27/2017	11:34
15	I1527-29	1	Solid	20.02	20	7.86	01/27/2017	11:35
16	CCV2	1	Water			12.02	01/27/2017	11:36
17	I1527-30	1	Solid	20.02	20	7.72	01/27/2017	11:37
18	I1528-21	1	Solid	20.04	20	7.81	01/27/2017	11:39
19	I1528-22	1	Solid	20.05	20	8.63	01/27/2017	11:40
20	I1528-23	1	Solid	20.01	20	7.87	01/27/2017	11:41
21	I1528-24	1	Solid	20.03	20	7.61	01/27/2017	11:42

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
22	I1528-25	1	Solid	20.02	20	7.76	01/27/2017	11:43
23	I1528-26	1	Solid	20.04	20	7.63	01/27/2017	11:44
24	I1528-27	1	Solid	20.00	20	6.90	01/27/2017	11:45
25	I1528-28	1	Solid	20.02	20	6.46	01/27/2017	11:46
26	I1528-29	1	Solid	20.02	20	6.83	01/27/2017	11:47
27	CCV3	1	Water			2.02	01/27/2017	11:48
28	I1528-30	1	Solid	20.02	20	7.88	01/27/2017	11:49
29	CCV4	1	Water			12.01	01/27/2017	11:50

LB8538

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I1527 WorkList ID : 94920

Date : 1/27/2017 11:06:05 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1527-21	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-01-COMP	01/23/2017	9045C
	Solid	I1527-22	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-02-COMP	01/23/2017	9045C
	Solid	I1527-23	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-04-COMP	01/23/2017	9045C
	Solid	I1527-24	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-05-COMP	01/23/2017	9045C
	Solid	I1527-25	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-06-COMP	01/23/2017	9045C
	Solid	I1527-26	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-07-COMP	01/23/2017	9045C
	Solid	I1527-27	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-10-COMP	01/24/2017	9045C
	Solid	I1527-28	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-11-COMP	01/24/2017	9045C
	Solid	I1527-29	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-12-COMP	01/24/2017	9045C
	Solid	I1527-30	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-13-COMP	01/24/2017	9045C
	Solid	I1528-21	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-14-COMP	01/24/2017	9045C
	Solid	I1528-22	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-15-COMP	01/24/2017	9045C
	Solid	I1528-23	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-16-COMP	01/24/2017	9045C
	Solid	I1528-24	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-17-COMP	01/24/2017	9045C
	Solid	I1528-25	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-18-COMP	01/24/2017	9045C
	Solid	I1528-26	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-19-COMP	01/25/2017	9045C
	Solid	I1528-27	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-20-COMP	01/25/2017	9045C
	Solid	I1528-28	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-21-COMP	01/25/2017	9045C
	Solid	I1528-29	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-23-COMP	01/26/2017	9045C
	Solid	I1528-30	Corrosivity	Cool 4 deg C	LIRO01	E21	SB-24-COMP	01/26/2017	9045C

Date/Time 01-27-16 11:06 AM
Received by: JP
Relinquished by: JP

Date/Time 01-27-16 5:30 PM
Received by: JP
Relinquished by: JP