


7470A

LB85549
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37674



A= 0.0000e+000

B= 2.6652e-004

C= -1.1388e-002

Rho= 0.9999490

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	14	0.000	14				
0.2	0.200	0.185	-0.015	736	0.0 %	736				
2.5	2.500	2.536	0.036	9557	0.0 %	9557				
5.0	5.000	5.029	0.029	18911	0.0 %	18911				
7.5	7.500	7.429	-0.071	27916	0.0 %	27916				
10.0	10.000	10.030	0.030	37674	0.0 %	37674				

LB85549 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Jan 2017 14:18:33

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	14	-	0.0000	7470A	PPB	27 Jan 2017 14:22:44	S	Std
0.2 - 1	50.2	736	-	0.2000	7470A	PPB	27 Jan 2017 14:24:52	S	Std
2.5 - 1	52.5	9557	-	2.5000	7470A	PPB	27 Jan 2017 14:27:01	S	Std
5.0 - 1	55	18911	-	5.0000	7470A	PPB	27 Jan 2017 14:29:09	S	Std
7.5 - 1	57.5	27916	-	7.5000	7470A	PPB	27 Jan 2017 14:31:17	S	Std
10.0 - 1	510	37674	-	10.0000	7470A	PPB	27 Jan 2017 14:33:24	S	Std
ICV09 - 1	ICV09	14859	3.9488	-	7470A	PPB	27 Jan 2017 14:36:07	U	SMPL
ICB09 - 1	ICB09	-15	-0.0154	-	7470A	PPB	27 Jan 2017 14:38:13	U	SMPL
CCV23 - 1	CCV23	18937	5.0357	-	7470A	PPB	27 Jan 2017 14:40:20	U	SMPL
CCB23 - 1	CCB23	-11	-0.0143	-	7470A	PPB	27 Jan 2017 14:42:27	U	SMPL
CRA - 1	CRA	737	0.1850	-	7470A	PPB	27 Jan 2017 14:44:33	U	SMPL
HighStd - 1	HighStd	37794	10.0615	-	7470A	PPB	27 Jan 2017 14:46:39	U	SMPL
ChkStd - 1	ChkStd	26752	7.1186	-	7470A	PPB	27 Jan 2017 14:48:46	U	SMPL
PB96375BL - 1	PBW	-16	-0.0157	-	7470A	PPB	27 Jan 2017 14:50:53	U	SMPL
PB96375BS - 1	LCSW	16882	4.4880	-	7470A	PPB	27 Jan 2017 14:55:51	U	SMPL
I1448-01 - 1	CHA-GW-01	93	0.0134	-	7470A	PPB	27 Jan 2017 14:57:58	U	SMPL
I1448-01DUP - 1	CHA-GW-01DUP	97	0.0145	-	7470A	PPB	27 Jan 2017 15:00:05	U	SMPL
I1448-01MS - 1	CHA-GW-01MS	14884	3.9555	-	7470A	PPB	27 Jan 2017 15:04:38	U	SMPL
I1448-01MSD - 1	CHA-GW-01MSD	14914	3.9635	-	7470A	PPB	27 Jan 2017 15:06:46	U	SMPL
I1448-05 - 1	CHA-FB012417	-17	-0.0159	-	7470A	PPB	27 Jan 2017 15:08:53	U	SMPL
CCV24 - 1	CCV24	18952	5.0397	-	7470A	PPB	27 Jan 2017 15:10:59	U	SMPL
CCB24 - 1	CCB24	-22	-0.0173	-	7470A	PPB	27 Jan 2017 15:13:06	U	SMPL
PB96376BL - 1	PBW	17	-0.0069	-	7470A	PPB	27 Jan 2017 15:15:13	U	SMPL
PB96376BS - 1	LCSW	15992	4.2508	-	7470A	PPB	27 Jan 2017 15:17:19	U	SMPL
PB96376TB - 1	PB96325TB	-17	-0.0159	-	7470A	PPB	27 Jan 2017 15:19:26	U	SMPL
I1448-09 - 1	CHA-WCS-012317-1(0-70)	9	-0.0090	-	7470A	PPB	27 Jan 2017 15:21:33	U	SMPL
I1448-09DUP - 1	CHA-WCS-012317-1(0-70)DUP	11	-0.0085	-	7470A	PPB	27 Jan 2017 15:23:41	U	SMPL
I1448-09MS - 1	CHA-WCS-012317-1(0-70)MS	16838	4.4763	-	7470A	PPB	27 Jan 2017 15:29:28	U	SMPL
I1448-09MSD - 1	CHA-WCS-012317-1(0-70)MSD	16911	4.4957	-	7470A	PPB	27 Jan 2017 15:31:35	U	SMPL
I1448-11 - 1	CHA-WCS-012417-1(0-4)	30	-0.0034	-	7470A	PPB	27 Jan 2017 15:33:43	U	SMPL
CCV25 - 1	CCV25	19136	5.0888	-	7470A	PPB	27 Jan 2017 15:35:52	U	SMPL
CCB25 - 1	CCB25	-23	-0.0175	-	7470A	PPB	27 Jan 2017 15:37:59	U	SMPL
I1448-01LX5 - 1		27	-0.0042	-	7470A	PPB	27 Jan 2017 15:40:06	U	SMPL
I1448-01A - 1		15492	4.1176	-	7470A	PPB	27 Jan 2017 15:42:13	U	SMPL
I1448-09LX5 - 1		11	-0.0085	-	7470A	PPB	27 Jan 2017 15:44:20	U	SMPL
I1448-09A - 1		15595	4.1450	-	7470A	PPB	27 Jan 2017 15:46:27	U	SMPL
CCV26 - 1	CCV26	19280	5.1271	-	7470A	PPB	27 Jan 2017 15:48:34	U	SMPL
CCB26 - 1	CCB26	-14	-0.0151	-	7470A	PPB	27 Jan 2017 15:50:41	U	SMPL