

Analytical Summary ReportAnalysis Method: 9038ANALYST: riteshParameter: SulfateSUPERVISOR REVIEW BY: apatelRun Number: LB85550BalanceID: WC SC-3

Reagent/Standard	Lot/Log #
sulfate ICV 20ppm	WP52435
Sulfate LCS std, 20ppm	WP52437
sulfate 0.0ppm	WP52426
sulfate 5ppm	WP52427
sulfate 10ppm	WP52428
sulfate 20ppm	WP52430
sulfate 30ppm	WP52432
sulfate 40ppm	WP52434
sulfate CCV 20ppm	WP52438
Sulfate std, 15ppm	WP52429
Sulfate std, 35ppm	WP52433
Sulfate std, 25ppm	WP52431
Barium Chloride, 500 gms	W1336
conditioning reagent	WP52505

Intercept: 0.1317Slope: 3.0079Regression: 0.999856

Seq	Lab ID	True Val (mg/l)	DF	MATRIX	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	Water	0.134	0.485	0.351	01/25/2017	11:01
2	CAL2	5	1	Water	0.150	14.300	14.150	01/25/2017	11:06
3	CAL3	10	1	Water	0.145	30.400	30.255	01/25/2017	11:11
4	CAL4	15	1	Water	0.139	46.700	46.561	01/25/2017	11:16
5	CAL5	20	1	Water	0.180	60.100	59.920	01/25/2017	11:21
6	CAL6	25	1	Water	0.160	75.100	74.940	01/25/2017	11:26
7	CAL7	30	1	Water	0.168	90.900	90.732	01/25/2017	11:31
8	CAL8	35	1	Water	0.198	106.100	105.902	01/25/2017	11:36
9	CAL9	40	1	Water	0.204	120.000	119.796	01/25/2017	11:41

Analysis Method: 9038
Parameter: Sulfate
Run Number: LB85550

ANALYST: ritesh
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	MATRIX	Initial Reading	Final Reading	Difference	Result (mg/L)	AnalDate	Anal Time
1	ICV			20	1	Water	0.298	62.900	62.602	20.769	01/25/2017	11:46
2	ICB				1	Water	0.108	0.285	0.177	0.015	01/25/2017	11:51
3	CCV1			20	1	Water	0.268	61.700	61.432	20.380	01/25/2017	11:56
4	CCB1				1	Water	0.111	0.301	0.190	0.019	01/25/2017	12:01
5	LB85550BL	5.00	50		1	Solid	0.095	0.228	0.133	0.000	01/25/2017	12:06
6	LB85550BS	5.00	50	20	1	Solid	0.368	63.200	62.832	20.845	01/25/2017	12:11
7	I1272-05	5.00	50		1	Solid	150.000	495.000	345.000	114.654	01/25/2017	12:16
8	I1272-05	5.00	50		4	Solid	77.500	161.000	83.500	27.716	01/25/2017	12:21
9	CCV2			20	1	Water	0.296	62.100	61.804	20.503	01/25/2017	12:26
10	CCB2				1	Water	0.120	0.298	0.178	0.015	01/25/2017	12:31