


245.1

LB 85552
INSTRUMENT ID: CV2

Linear

 μ Abs.:

39129



A= 0.0000e+000

B= 2.5543e-004

C= -9.3608e-003

Rho= 0.9999784

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	-4	0.000	-4				
0.2	0.200	0.188	-0.012	774	0.0 %	774				
2.5	2.500	2.540	0.040	9982	0.0 %	9982				
5.0	5.000	4.971	-0.029	19499	0.0 %	19499				
7.5	7.500	7.525	0.025	29496	0.0 %	29496				
10.0	10.000	9.985	-0.015	39129	0.0 %	39129				

LB85552 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Jan 2017 16:03:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-4	-	0.0000	245.1	PPB	27 Jan 2017 16:17:44	S	Std
0.2 - 1	50.2	774	-	0.2000	245.1	PPB	27 Jan 2017 16:19:52	S	Std
2.5 - 1	52.5	9982	-	2.5000	245.1	PPB	27 Jan 2017 16:22:00	S	Std
5.0 - 1	55	19499	-	5.0000	245.1	PPB	27 Jan 2017 16:24:09	S	Std
7.5 - 1	57.5	29496	-	7.5000	245.1	PPB	27 Jan 2017 16:26:17	S	Std
10.0 - 1	510	39129	-	10.0000	245.1	PPB	27 Jan 2017 16:33:37	S	Std
ICV10 - 1	ICV10	16321	4.1596	-245.1		PPB	27 Jan 2017 16:40:30	U	SMPL
ICB10 - 1	ICB10	-19	-0.0142	-245.1		PPB	27 Jan 2017 16:42:36	U	SMPL
CCV27 - 1	CCV27	19797	5.0474	-245.1		PPB	27 Jan 2017 16:44:43	U	SMPL
CCB27 - 1	CCB27	-34	-0.0180	-245.1		PPB	27 Jan 2017 16:46:49	U	SMPL
CRA - 1	CRA	766	0.1863	-245.1		PPB	27 Jan 2017 16:48:56	U	SMPL
HighStd - 1	HighStd	39558	10.0950	-245.1		PPB	27 Jan 2017 16:51:03	U	SMPL
ChkStd - 1	ChkStd	25075	6.3956	-245.1		PPB	27 Jan 2017 16:53:10	U	SMPL
PB96374BL - 1	PBW	-33	-0.0178	-245.1		PPB	27 Jan 2017 16:55:16	U	SMPL
PB96374BS - 1	LCSW	15048	3.8344	-245.1		PPB	27 Jan 2017 16:57:23	U	SMPL
I1417-01 - 1	012017-DSDC-E-D-M	48	0.0029	-245.1		PPB	27 Jan 2017 16:59:31	U	SMPL
I1417-02 - 1	012017-DS-E-U-M	60	0.0060	-245.1		PPB	27 Jan 2017 17:01:39	U	SMPL
I1417-02DUP - 1	012017-DS-E-U-MDUP	66	0.0075	-245.1		PPB	27 Jan 2017 17:03:47	U	SMPL
I1417-02MS - 1	012017-DS-E-U-MMS	16205	4.1299	-245.1		PPB	27 Jan 2017 17:05:56	U	SMPL
I1417-02MSD - 1	012017-DS-E-U-MMSD	16274	4.1476	-245.1		PPB	27 Jan 2017 17:08:03	U	SMPL
CCV28 - 1	CCV28	20351	5.1890	-245.1		PPB	27 Jan 2017 17:10:10	U	SMPL
CCB28 - 1	CCB28	-28	-0.0165	-245.1		PPB	27 Jan 2017 17:12:16	U	SMPL
I1417-02LX5 - 1		0	-0.0094	-245.1		PPB	27 Jan 2017 17:14:23	U	SMPL
I1417-02A - 1		14907	3.7984	-245.1		PPB	27 Jan 2017 17:18:53	U	SMPL
CCV29 - 1	CCV29	20439	5.2114	-245.1		PPB	27 Jan 2017 17:20:59	U	SMPL
CCB29 - 1	CCB29	-35	-0.0183	-245.1		PPB	27 Jan 2017 17:23:06	U	SMPL