



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB85565

Reagent/Standard	Lot/Loq #
TURBIDITY LOD STD, 0.5NTU	WP52571
TURBIDITY LOQ STD, 2.0NTU	WP52572
TURBIDITY CCV 10 NTU	WP52566
Turbidity std, 40NTU	WP52551
turbidity std 5 NTU	WP52553
turbidity std 20 NTU	WP52555
Turbidity std 1 NTU	WP52552
turbidity std 0 NTU	WP52547
turbidity std 10 NTU	WP52554

Intercept: 0.0625

Slope: 0.9678

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.078	01/30/2017	08:49
2	CAL2	1	1	1.050	01/30/2017	08:51
3	CAL3	5	1	4.950	01/30/2017	08:53
4	CAL4	10	1	9.850	01/30/2017	08:55
5	CAL5	20	1	19.100	01/30/2017	08:57
6	CAL6	40	1	38.900	01/30/2017	08:59



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB85565

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.820	10.082	01/30/2017	09:01
2	ICB		1	0.075	0.013	01/30/2017	09:03
3	CCV1	10	1	9.850	10.113	01/30/2017	09:05
4	CCB1		1	0.074	0.012	01/30/2017	09:07
5	LB85565BL		1	0.075	0.013	01/30/2017	09:09
6	I1092-05		1	0.600	0.555	01/30/2017	09:11
7	I1092-06		1	2.010	2.012	01/30/2017	09:13
8	I1272-10		1	10.400	10.681	01/30/2017	09:15
9	I1307-02		1	3.020	3.056	01/30/2017	09:17
10	CCV2	10	1	9.850	10.113	01/30/2017	09:19
11	CCB2		1	0.075	0.013	01/30/2017	09:21