



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB85566

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP52566
Turbidity std, 40NTU	WP52551
turbidity std 5 NTU	WP52553
turbidity std 20 NTU	WP52555
Turbidity std 1 NTU	WP52552
turbidity std 0 NTU	WP52547
turbidity std 10 NTU	WP52554
10 NTU Standard 500 ml	W2152

Intercept: 0.0616

Slope: 0.9678

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.078	01/30/2017	08:49
2	CAL2	1	1	1.050	01/30/2017	08:51
3	CAL3	5	1	4.950	01/30/2017	08:53
4	CAL4	10	1	9.850	01/30/2017	08:55
5	CAL5	20	1	19.100	01/30/2017	08:57
6	CAL6	40	1	38.900	01/30/2017	08:59



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB85566

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.820	10.083	01/30/2017	09:01
2	ICB		1	0.075	0.014	02/02/2017	09:03
3	CCV1	10	1	9.849	10.113	02/02/2017	09:23
4	CCB1		1	0.076	0.015	01/30/2017	09:25
5	LB85566BL		1	0.075	0.014	01/30/2017	09:27
6	I1092-03		1	0.558	0.513	01/30/2017	09:29
7	I1092-04		1	2.000	2.003	01/30/2017	09:31
8	I1272-09		1	10.410	10.693	01/30/2017	09:33
9	I1307-01		1	3.010	3.046	01/30/2017	09:35
10	CCV2	10	1	9.875	10.140	01/30/2017	09:37
11	CCB2		1	0.076	0.015	01/30/2017	09:39