


7471A

LB 8556 8
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44624



A= 0.0000e+000

B= 2.2642e-004

C= -6.6602e-004

Rho= 0.9997465

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	23	0.000	23				
0.2	0.200	0.192	-0.008	853	0.0 %	853				
2.5	2.500	2.596	0.096	11468	0.0 %	11468				
5.0	5.000	4.936	-0.064	21804	0.0 %	21804				
7.5	7.500	7.368	-0.132	32544	0.0 %	32544				
10.0	10.000	10.103	0.103	44624	0.0 %	44624				

LB85568 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 30 Jan 2017 11:31:17					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	7471A	PPB	30 Jan 2017 13:39:56	S	Std
0.2 - 1	50-2	853	-	0.2000	7471A	PPB	30 Jan 2017 13:42:13	S	Std
2.5 - 1	52-5	11468	-	2.5000	7471A	PPB	30 Jan 2017 13:44:30	S	Std
5.0 - 1	55	21804	-	5.0000	7471A	PPB	30 Jan 2017 13:46:48	S	Std
7.5 - 1	57-5	32544	-	7.5000	7471A	PPB	30 Jan 2017 13:49:05	S	Std
10.0 - 1	510	44624	-	10.0000	7471A	PPB	30 Jan 2017 13:54:03	S	Std
ICV11 - 1	ICV11	16670	3.7737	-7471A	PPB	PPB	30 Jan 2017 13:56:47	U	SMPL
ICB11 - 1	ICB11	-34	-0.0084	-7471A	PPB	PPB	30 Jan 2017 13:59:02	U	SMPL
CCV30 - 1	CCV30	22046	4.9910	-7471A	PPB	PPB	30 Jan 2017 14:01:18	U	SMPL
CCB30 - 1	CCB30	-21	-0.0054	-7471A	PPB	PPB	30 Jan 2017 14:03:34	U	SMPL
CRA - 1	CRA	843	0.1902	-7471A	PPB	PPB	30 Jan 2017 14:05:50	U	SMPL
HighStd - 1	HighStd	45334	10.2638	-7471A	PPB	PPB	30 Jan 2017 14:08:05	U	SMPL
ChkStd - 1	ChkStd	30208	6.8390	-7471A	PPB	PPB	30 Jan 2017 14:10:22	U	SMPL
PB96381BL - 1	PBS	-50	-0.0120	-7471A	PPB	PPB	30 Jan 2017 14:12:38	U	SMPL
PB96381BS - 1	LCSS	16687	3.7776	-7471A	PPB	PPB	30 Jan 2017 14:14:54	U	SMPL
I1386-01 - 1	CHA-SB-01-0-2	3330	0.7533	-7471A	PPB	PPB	30 Jan 2017 14:17:11	U	SMPL
I1386-01DUP - 1	CHA-SB-01-0-2DUP	3942	0.8919	-7471A	PPB	PPB	30 Jan 2017 14:19:29	U	SMPL
I1386-02 - 1	I1386-01MS	21351	4.8336	-7471A	PPB	PPB	30 Jan 2017 14:21:46	U	SMPL
I1386-03 - 1	I1386-01MSD	20288	4.5929	-7471A	PPB	PPB	30 Jan 2017 14:24:04	U	SMPL
I1386-04 - 1	CHA-SB-101-0-2	3688	0.8344	-7471A	PPB	PPB	30 Jan 2017 14:26:20	U	SMPL
CCV31 - 1	CCV31	21968	4.9733	-7471A	PPB	PPB	30 Jan 2017 14:28:36	U	SMPL
CCB31 - 1	CCB31	-44	-0.0106	-7471A	PPB	PPB	30 Jan 2017 14:30:53	U	SMPL
I1386-05 - 1	CHA-SB-01-9-11	77	0.0168	-7471A	PPB	PPB	30 Jan 2017 14:33:09	U	SMPL
I1408-01 - 1	TILCON-STONEDUST-SE	731	0.1648	-7471A	PPB	PPB	30 Jan 2017 14:35:25	U	SMPL
I1413-01 - 1	TILCON-STONEDUST-BF	447	0.1005	-7471A	PPB	PPB	30 Jan 2017 14:37:40	U	SMPL
I1448-06 - 1	CHA-SB-02-2-4	578	0.1302	-7471A	PPB	PPB	30 Jan 2017 14:39:56	U	SMPL
I1448-07 - 1	CHA-SB-03-2-4	315	0.0707	-7471A	PPB	PPB	30 Jan 2017 14:42:12	U	SMPL
I1448-08 - 1	CHA-BL-01-0-2	243	0.0544	-7471A	PPB	PPB	30 Jan 2017 14:44:29	U	SMPL
I1386-01LX5 - 1		616	0.1388	-7471A	PPB	PPB	30 Jan 2017 14:46:47	U	SMPL
I1386-01A - 1		20547	4.6516	-7471A	PPB	PPB	30 Jan 2017 14:49:04	U	SMPL
CCV32 - 1	CCV32	21898	4.9574	-7471A	PPB	PPB	30 Jan 2017 14:51:22	U	SMPL
CCB32 - 1	CCB32	-43	-0.0104	-7471A	PPB	PPB	30 Jan 2017 14:53:38	U	SMPL