

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/31/17

Analyst : AP

Data File : LB85579.csv

Approved By :

Approved Date :

Worksheet # :

M300-D - Inorganic Arsenic - U

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Bromide											
STD1	STD1			W	1/27/17						
Bromide		PASS		0.000	0.000					mg/L	
STD2	STD2			W	1/27/17						
Bromide		PASS		0.025	0.025					mg/L	
STD3	STD3			W	1/27/17						
Bromide		PASS		0.048	0.048					mg/L	
STD4	STD4			W	1/27/17						
Bromide		PASS		0.098	0.098					mg/L	
STD5	STD5			W	1/27/17						
Bromide		PASS		0.203	0.203					mg/L	
STD6	STD6			W	1/27/17						
Bromide		PASS		0.504	0.504					mg/L	
STD7	STD7			W	1/27/17						
Bromide		PASS		0.747	0.747					mg/L	
STD8	STD8			W	1/27/17						
Bromide		PASS		1.000	1.00					mg/L	
CCV1	CCV1			W	1/27/17						
Bromide		PASS		0.505	0.51	102.0	90	110		mg/L	
CCB1	CCB1			W	1/27/17						
Bromide		PASS		0.000	0.000			+/-0.0130		mg/L	
ICV1	ICV1			W	1/27/17						
Bromide		PASS		0.503	0.50	100.0	90	110		mg/L	
ICB1	ICB1			W	1/27/17						
Bromide		PASS		0.000	0.000			+/-0.0130		mg/L	
LB85579BLW	LB85579BLW			W	1/27/17						
Bromide		PASS		0.000	0.000			+/-0.0130		mg/L	
LB85579BSW	LB85579BSW			W	1/27/17						
Bromide		PASS		0.503	0.50	100.0	90	110		mg/L	
I1307-06	PT-IDB1-WS		1	W	1/27/17						
Bromide		PASS		0.051	0.051					mg/L	
CCV1	CCV1			W	1/27/17						
Bromide		PASS		0.507	0.51	102.0	90	110		mg/L	
CCB1	CCB1			W	1/27/17						
Bromide		PASS		0.000	0.000			+/-0.0130		mg/L	
LB85579BLW2	LB85579BLW2			W	1/27/17						
Bromide		PASS		0.000	0.000			+/-0.0130		mg/L	
LB85579BSW2	LB85579BSW2			W	1/27/17						
Bromide		PASS		0.505	0.51	102.0	90	110		mg/L	
I1002-08	MDL-WATER-1		1	W	1/27/17						
Bromide		PASS		0.025	0.025					mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/31/17
Analyst : AP
Data File : LB85579.csv

Approved By : 
Approved Date : 1/31/17
Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Bromide											
I1002-09	MDL-WATER-2		1	W	1/27/17						
Bromide	PASS	0.023		0.023						mg/L	
I1002-10	MDL-WATER-3		1	W	1/27/17						
Bromide	PASS	0.027		0.027						mg/L	
I1002-11	MDL-WATER-4		1	W	1/27/17						
Bromide	PASS	0.025		0.025						mg/L	
I1002-12	MDL-WATER-5		1	W	1/27/17						
Bromide	PASS	0.024		0.024						mg/L	
I1002-13	MDL-WATER-6		1	W	1/27/17						
Bromide	PASS	0.023		0.023						mg/L	
I1002-14	MDL-WATER-7		1	W	1/27/17						
Bromide	PASS	0.025		0.025						mg/L	
I1002-14	MDL-WATER-7		1	W	1/27/17						
Bromide	PASS	0.022		0.022						mg/L	
CCV2	CCV2			W	1/28/17						
Bromide	PASS	0.503		0.50	100.0	90	110			mg/L	
CCB2	CCB2			W	1/28/17						
Bromide	PASS	0.000		0.000			+/-0.0130			mg/L	