

Analysis Method: SM2510 BANALYST: AMANDEEPParameter: ConductanceSUPERVISOR REVIEW BY: apatelRun Number: LB85587

| Reagent/Standard                | Lot/Log # |
|---------------------------------|-----------|
| conductivity standar (CAL/LCSW) | WP49968   |

| Seq | Lab ID    | True Value<br>(mg/L) | Dilution | Temperature<br>°C | Reading<br>(umho/cm) | Result<br>(umho/cm) | Anal Date  | Anal Time |
|-----|-----------|----------------------|----------|-------------------|----------------------|---------------------|------------|-----------|
| 1   | LB85587BL |                      | 1        | 21.3              | 0.92                 | 0.99                | 01/30/2017 | 16:01     |
| 2   | LB85587BS | 1413                 | 1        | 21.3              | 1275.00              | 1371.96             | 01/30/2017 | 16:03     |
| 3   | I1272-13  |                      | 1        | 21.3              | 675.00               | 726.33              | 01/30/2017 | 16:05     |

Result =                     (Reading)  
1 + 0.0191 (Temperature - 25)

|