

Analysis Method: 120.1
Parameter: Conductance
Run Number: LB85588

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Reagent/Standard	Lot/Log #
conductivity standar (CAL/LCSW)	WP49968

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB85588BL		1	21.3	0.91	0.98	01/30/2017	16:07
2	LB85588BS	1413	1	21.3	1275.00	1371.96	01/30/2017	16:09
3	I1272-14		1	21.3	673.00	724.18	01/30/2017	16:11

Result = (Reading)
1 + 0.0191 (Temperature - 25)

|