

Analysis Method: 9050A

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB85589

Reagent/Standard	Lot/Log #
conductivity standar (CAL/LCSW)	WP49968

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB85589BL		1	21.3	0.90	0.97	01/30/2017	16:13
2	LB85589BS	1413	1	21.3	1275.00	1371.96	01/30/2017	16:15
3	I1272-15		1	21.3	674.00	725.25	01/30/2017	16:17

Result = (Reading)
1 + 0.0191 (Temperature - 25)

|