



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85591

Temp.: 22.5 °C

BalanceID:

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/30/2017	14:00
2	CAL2	1	Water			7.02	01/30/2017	14:01
3	CAL3	1	Water			10.04	01/30/2017	14:02
4	ICV	1	Water			7.01	01/30/2017	14:03
5	CCV1	1	Water			2.02	01/30/2017	14:05
6	I1572-01	1	Solid	20.02	20	11.0	01/30/2017	14:06
7	I1588-04	1	Solid	20.04	20	8.69	01/30/2017	14:07
8	I1593-10	1	Solid	20.01	20	7.86	01/30/2017	14:08
9	I1593-11	1	Solid	20.02	20	7.88	01/30/2017	14:09
10	I1593-12	1	Solid	20.01	20	7.69	01/30/2017	14:10
11	I1593-12DUP	1	Solid	20.03	20	7.70	01/30/2017	14:12
12	CCV2	1	Water			12.01	01/30/2017	14:13

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 11588

WorkList ID : 95122

Date : 1/30/2017 1:19:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/01/2017	Solid	I1572-01	Corrosivity	Cool 4 deg C	DIST01	F32	COMPOSITE-1	01/27/2017	9045C
02/01/2017	Solid	I1588-04	Corrosivity	Cool 4 deg C	ENTA04	F31	WC-B15-B21-001	01/26/2017	9045C
	Solid	I1593-10	Corrosivity	Cool 4 deg C	STV101	F32	B2-6	01/25/2017	9045C
	Solid	I1593-11	Corrosivity	Cool 4 deg C	STV101	F32	B2-7	01/26/2017	9045C
	Solid	I1593-12	Corrosivity	Cool 4 deg C	STV101	F32	B2-8	01/24/2017	9045C

Date/Time 01-30-17 1:19 PM
 Received by: JP
 Relinquished by: CQ

Date/Time 01-20-17 5:00 PM
 Received by: UP
 Relinquished by: JP