

7470A


LB 85596

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37191



A= 0.0000e+000

B= 2.6940e-004

C= 1.6956e-002

Rho= 0.9999584

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	27	0.000	27				
0.2	0.200	0.221	0.021	758	0.0 %	758				
2.5	2.500	2.485	-0.015	9160	0.0 %	9160				
5.0	5.000	4.935	-0.065	18256	0.0 %	18256				
7.5	7.500	7.499	-0.001	27771	0.0 %	27771				
10.0	10.000	10.036	0.036	37191	0.0 %	37191				

LB85596 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Jan 2017 11:33:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.00007470A	PPB		31 Jan 2017 11:38:07	S	Std
0.2 - 1	50.2	758	-	0.20007470A	PPB		31 Jan 2017 11:40:16	S	Std
2.5 - 1	52.5	9160	-	2.50007470A	PPB		31 Jan 2017 11:42:24	S	Std
5.0 - 1	55	18256	-	5.00007470A	PPB		31 Jan 2017 11:44:33	S	Std
7.5 - 1	57.5	27771	-	7.50007470A	PPB		31 Jan 2017 11:51:16	S	Std
10.0 - 1	510	37191	-	10.00007470A	PPB		31 Jan 2017 11:53:23	S	Std
ICV12 - 1	ICV12	14116	3.8198	-7470A	PPB		31 Jan 2017 11:56:02	U	SMPL
ICB12 - 1	ICB12	0	0.0170	-7470A	PPB		31 Jan 2017 11:58:08	U	SMPL
CCV36 - 1	CCV36	18401	4.9742	-7470A	PPB		31 Jan 2017 12:00:14	U	SMPL
CCB36 - 1	CCB36	-7	0.0151	-7470A	PPB		31 Jan 2017 12:02:21	U	SMPL
CRA - 1	CRA	747	0.2182	-7470A	PPB		31 Jan 2017 12:04:27	U	SMPL
HighStd - 1	HighStd	37801	10.2006	-7470A	PPB		31 Jan 2017 12:06:34	U	SMPL
ChkStd - 1	ChkStd	26297	7.1014	-7470A	PPB		31 Jan 2017 12:08:41	U	SMPL
PB96542BL - 1	PBW	-27	0.0097	-7470A	PPB		31 Jan 2017 12:10:48	U	SMPL
PB96542BS - 1	LCSW	14335	3.8788	-7470A	PPB		31 Jan 2017 12:12:55	U	SMPL
PB96542TB - 1	PB96399TB	-29	0.0091	-7470A	PPB		31 Jan 2017 12:15:03	U	SMPL
I1527-21 - 1	SB-01-COMP	-6	0.0153	-7470A	PPB		31 Jan 2017 12:17:11	U	SMPL
I1527-22 - 1	SB-02-COMP	1	0.0172	-7470A	PPB		31 Jan 2017 12:19:20	U	SMPL
I1527-23 - 1	SB-04-COMP	2	0.0175	-7470A	PPB		31 Jan 2017 12:21:28	U	SMPL
I1527-24 - 1	SB-05-COMP	-1	0.0167	-7470A	PPB		31 Jan 2017 12:23:35	U	SMPL
CCV37 - 1	CCV37	18438	4.9842	-7470A	PPB		31 Jan 2017 12:25:42	U	SMPL
CCB37 - 1	CCB37	-3	0.0161	-7470A	PPB		31 Jan 2017 12:27:48	U	SMPL
I1527-25 - 1	SB-06-COMP	-8	0.0148	-7470A	PPB		31 Jan 2017 12:29:55	U	SMPL
I1527-26 - 1	SB-07-COMP	12	0.0202	-7470A	PPB		31 Jan 2017 12:32:02	U	SMPL
I1527-27 - 1	SB-10-COMP	17	0.0215	-7470A	PPB		31 Jan 2017 12:34:09	U	SMPL
I1527-28 - 1	SB-11-COMP	9	0.0194	-7470A	PPB		31 Jan 2017 12:36:16	U	SMPL
I1527-29 - 1	SB-12-COMP	10	0.0197	-7470A	PPB		31 Jan 2017 12:38:23	U	SMPL
I1527-30 - 1	SB-13-COMP	2	0.0175	-7470A	PPB		31 Jan 2017 12:40:31	U	SMPL
I1528-21 - 1	SB-14-COMP	-13	0.0135	-7470A	PPB		31 Jan 2017 12:42:39	U	SMPL
I1528-22 - 1	SB-15-COMP	0	0.0170	-7470A	PPB		31 Jan 2017 12:44:47	U	SMPL
I1528-23 - 1	SB-16-COMP	0	0.0170	-7470A	PPB		31 Jan 2017 12:46:56	U	SMPL
I1528-24 - 1	SB-17-COMP	2	0.0175	-7470A	PPB		31 Jan 2017 12:49:02	U	SMPL
CCV38 - 1	CCV38	18087	4.8896	-7470A	PPB		31 Jan 2017 12:51:09	U	SMPL
CCB38 - 1	CCB38	-7	0.0151	-7470A	PPB		31 Jan 2017 12:53:16	U	SMPL
I1528-25 - 1	SB-18-COMP	-1	0.0167	-7470A	PPB		31 Jan 2017 12:55:23	U	SMPL
I1528-26 - 1	SB-19-COMP	-6	0.0153	-7470A	PPB		31 Jan 2017 12:57:30	U	SMPL
I1528-27 - 1	SB-20-COMP	6	0.0186	-7470A	PPB		31 Jan 2017 12:59:36	U	SMPL
I1528-28 - 1	SB-21-COMP	12	0.0202	-7470A	PPB		31 Jan 2017 13:01:43	U	SMPL
I1528-29 - 1	SB-23-COMP	-5	0.0156	-7470A	PPB		31 Jan 2017 13:03:50	U	SMPL
I1528-30 - 1	SB-24-COMP	4	0.0180	-7470A	PPB		31 Jan 2017 13:05:58	U	SMPL
I1528-30DUP - 1	SB-24-COMPDUP	9	0.0194	-7470A	PPB		31 Jan 2017 13:08:06	U	SMPL
I1528-30MS - 1	SB-24-COMPMS	13734	3.7169	-7470A	PPB		31 Jan 2017 13:10:15	U	SMPL
I1528-30MSD - 1	SB-24-COMPMSD	13677	3.7016	-7470A	PPB		31 Jan 2017 13:12:23	U	SMPL
PB96543BL - 1	PBW	-22	0.0110	-7470A	PPB		31 Jan 2017 13:14:30	U	SMPL
CCV39 - 1	CCV39	18236	4.9298	-7470A	PPB		31 Jan 2017 13:16:37	U	SMPL
CCB39 - 1	CCB39	9	0.0194	-7470A	PPB		31 Jan 2017 13:18:45	U	SMPL
PB96543BS - 1	LCSW	13267	3.5911	-7470A	PPB		31 Jan 2017 13:20:52	U	SMPL
PB96543TB - 1	PB96400TB	-10	0.0143	-7470A	PPB		31 Jan 2017 13:22:59	U	SMPL
I1529-04 - 1	ELT-WCS012617(0-15)	1	0.0172	-7470A	PPB		31 Jan 2017 13:25:06	U	SMPL
I1529-05 - 1	ELT-WCS012617(15-60)	0	0.0170	-7470A	PPB		31 Jan 2017 13:27:13	U	SMPL
I1531-05 - 1	EDC-HPRR-S1C1	-12	0.0137	-7470A	PPB		31 Jan 2017 13:29:21	U	SMPL
I1531-06 - 1	EDC-HPRR-S2C2	5	0.0183	-7470A	PPB		31 Jan 2017 13:31:28	U	SMPL
I1531-07 - 1	EDC-HPRR-S3C3	12	0.0202	-7470A	PPB		31 Jan 2017 13:33:37	U	SMPL
I1531-08 - 1	EDC-HPRR-S4C4	-5	0.0156	-7470A	PPB		31 Jan 2017 13:35:45	U	SMPL
I1531-08DUP - 1	EDC-HPRR-S4C4DUP	-5	0.0156	-7470A	PPB		31 Jan 2017 13:37:54	U	SMPL
I1531-08MS - 1	EDC-HPRR-S4C4MS	13966	3.7794	-7470A	PPB		31 Jan 2017 13:40:01	U	SMPL
CCV40 - 1	CCV40	18024	4.8727	-7470A	PPB		31 Jan 2017 13:42:08	U	SMPL
CCB40 - 1	CCB40	19	0.0221	-7470A	PPB		31 Jan 2017 13:44:15	U	SMPL
I1531-08MSD - 1	EDC-HPRR-S4C4MSD	14035	3.7980	-7470A	PPB		31 Jan 2017 13:46:22	U	SMPL
I1528-30LX5 - 1		-15	0.0129	-7470A	PPB		31 Jan 2017 13:48:30	U	SMPL
I1528-30A - 1		15007	4.0599	-7470A	PPB		31 Jan 2017 13:50:37	U	SMPL
I1531-08LX5 - 1		-32	0.0083	-7470A	PPB		31 Jan 2017 13:52:44	U	SMPL
I1531-08A - 1		14899	4.0308	-7470A	PPB		31 Jan 2017 13:54:52	U	SMPL
CCV41 - 1	CCV41	18386	4.9702	-7470A	PPB		31 Jan 2017 13:56:59	U	SMPL
CCB41 - 1	CCB41	21	0.0226	-7470A	PPB		31 Jan 2017 13:59:07	U	SMPL