

**\*LB85611\***

**CHEMTECH**

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

**Analysis Method:** SM2320 B  
**Parameter:** Alkalinity  
**Run Number:** LB85611  
**Instrument:** TITROLINE  
**Analyst:** IB

MSM2320B - Alkalinity - 12

# Chemtech Consulting Group


## Analytical Review Report


Date Printed : 1/31/17

Analyst : IB

Data File : LB85611.MDB

Approved By : 

Approved Date : 

Worksheet # : 

| Lab Sample ID     | Client ID  |      | Raw Amt | Dil | Matrix     | A. Date | Prep | Analysis  |     |         |       | Line 1 |
|-------------------|------------|------|---------|-----|------------|---------|------|-----------|-----|---------|-------|--------|
| Parameter         |            |      | PPB     |     | Final Conc | %Rec    | LCL  | UCL       | RPD | Max RPD | Units | Line 2 |
| <b>Alkalinity</b> |            |      |         |     |            |         |      |           |     |         |       |        |
| LB85611BLW        | LB85611BLW |      |         |     | W          | 1/31/17 |      |           |     |         |       |        |
| Alkalinity        |            | PASS | 0.000   |     | 0.0000     |         |      | +/-0.4000 |     |         | mg/L  |        |
| LB85611BSW        | LB85611BSW |      |         |     | W          | 1/31/17 |      |           |     |         |       |        |
| Alkalinity        |            | PASS | 48.160  |     | 48.20      | 96.0    | 80   | 120       |     |         | mg/L  |        |
| I1272-22          | 0006       |      |         | 1   | W          | 1/31/17 |      |           |     |         |       |        |
| Alkalinity        |            | PASS | 146.880 |     | 147        |         |      |           |     |         | mg/L  |        |