

Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB85612

Reagent/Standard	Lot/Loq #
MBAS CAL STD, 0.5PPM	WP52603
MBAS CCV STD, 1.0PPM	WP52605
MBAS DAILY CHECK STD, 0.01PPM	WP52608
MBAS DAILY CHECK STD, 2.0PPM	WP52606
MBAS CAL STD 2 PPM	WP52621
MBAS CAL STD 1 PPM	WP52597
MBAS CAL STD 0.1 PPM	WP52600
mbas cal std 0.05 ppm	WP52599
MBAS CAL STD 0.01 PPM	WP52598
MBAS BLANK 0 PPM	WP52596
MBAS CAL STD 0.4 PPM	WP52602
MBAS CAL STD 0.2 PPM	WP52601
MBAS ICV-LCS std 1.0PPM	WP52604
MBAS LOD std, 0.005PPM	WP52609
MBAS LOQ std, 0.02PPM	WP52610

Intercept: -0.0008

Slope: 0.2148

Regression: 1

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Anal Date	Anal Time
1	CAL1	0	1	0.000	01/31/2017	11:25
2	CAL2	0.01	1	0.002	01/31/2017	11:25
3	CAL3	0.05	1	0.010	01/31/2017	11:26
4	CAL4	0.1	1	0.021	01/31/2017	11:26
5	CAL5	0.2	1	0.042	01/31/2017	11:27
6	CAL6	0.4	1	0.085	01/31/2017	11:27
7	CAL7	0.5	1	0.106	01/31/2017	11:28
8	CAL8	1	1	0.212	01/31/2017	11:28
9	CAL9	2	1	0.430	01/31/2017	11:29



Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB85612

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Result (mg/L)	AnalDate	Anal Time
1	ICV	1	1	0.209	0.977	01/31/2017	11:29
2	ICB		1	0.000	0.004	01/31/2017	11:30
3	CHK1	0.01	1	0.002	0.013	01/31/2017	11:31
4	CHK2	2	1	0.431	2.010	01/31/2017	11:31
5	CCV1	1	1	0.211	0.986	01/31/2017	11:31
6	CCB1		1	0.000	0.004	01/31/2017	11:32
7	LB85612BL		1	0.000	0.004	01/31/2017	11:32
8	LB85612BS	1	1	0.208	0.972	01/31/2017	11:33
9	I1092-03		1	0.001	0.008	01/31/2017	11:33
10	I1092-04		1	0.005	0.027	01/31/2017	11:34
11	I1272-26		1	0.186	0.870	01/31/2017	11:34
12	CCV2	1	1	0.212	0.991	01/31/2017	11:35
13	CCB2		1	0.000	0.004	01/31/2017	11:35