


7471A

LB 85614
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44729



A= 0.0000e+000

B= 2.2635e-004

C= -4.5940e-002

Rho= 0.9995662

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.048	-0.048	-8	0.000	-8				
0.2	0.200	0.164	-0.036	928	0.0 %	928				
2.5	2.500	2.676	0.176	12024	0.0 %	12024				
5.0	5.000	5.003	0.003	22308	0.0 %	22308				
7.5	7.500	7.326	-0.174	32569	0.0 %	32569				
10.0	10.000	10.078	0.078	44729	0.0 %	44729				

LB85614 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 31 Jan 2017 17:27:42

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-8	-	0.00007471A	PPB		31 Jan 2017 17:33:38	S	Std
0.2 - 1	50.2	928	-	0.20007471A	PPB		31 Jan 2017 17:35:55	S	Std
2.5 - 1	52.5	12024	-	2.50007471A	PPB		31 Jan 2017 17:38:13	S	Std
5.0 - 1	55	22308	-	5.00007471A	PPB		31 Jan 2017 17:40:31	S	Std
7.5 - 1	57.5	32569	-	7.50007471A	PPB		31 Jan 2017 17:42:48	S	Std
10.0 - 1	510	44729	-	10.00007471A	PPB		31 Jan 2017 17:47:50	S	Std
ICV13 - 1	ICV13	17689	3.9580	-7471A	PPB		31 Jan 2017 17:50:35	U	SMPL
ICB13 - 1	ICB13	-25	-0.0516	-7471A	PPB		31 Jan 2017 17:52:50	U	SMPL
CCV42 - 1	CCV42	23007	5.1617	-7471A	PPB		31 Jan 2017 17:55:06	U	SMPL
CCB42 - 1	CCB42	-31	-0.0530	-7471A	PPB		31 Jan 2017 17:57:22	U	SMPL
CRA - 1	CRA	917	0.1616	-7471A	PPB		31 Jan 2017 17:59:38	U	SMPL
HighStd - 1	HighStd	44628	10.0556	-7471A	PPB		31 Jan 2017 18:01:53	U	SMPL
ChkStd - 1	ChkStd	29087	6.5379	-7471A	PPB		31 Jan 2017 18:04:10	U	SMPL
PB96556BL - 1	PBS	-34	-0.0536	-7471A	PPB		31 Jan 2017 18:06:26	U	SMPL
I1529-01 - 1	ELT-SB-02-13-15	401	0.0448	-7471A	PPB		31 Jan 2017 18:10:59	U	SMPL
I1529-02 - 1	ELT-SB-01-58-60	52	-0.0342	-7471A	PPB		31 Jan 2017 18:13:17	U	SMPL
I1529-03 - 1	ELT-BL-0-2	823	0.1403	-7471A	PPB		31 Jan 2017 18:15:34	U	SMPL
I1531-01 - 1	EDC-HPRR-S1C1	5357	1.1666	-7471A	PPB		31 Jan 2017 18:17:52	U	SMPL
I1531-02 - 1	EDC-HPRR-S2C2	14411	3.2160	-7471A	PPB		31 Jan 2017 18:20:08	U	SMPL
CCV43 - 1	CCV43	22912	5.1402	-7471A	PPB		31 Jan 2017 18:22:24	U	SMPL
CCB43 - 1	CCB43	-44	-0.0559	-7471A	PPB		31 Jan 2017 18:24:39	U	SMPL
I1531-03 - 1	EDC-HPRR-S3C3	15197	3.3939	-7471A	PPB		31 Jan 2017 18:26:56	U	SMPL
I1531-04 - 1	EDC-HPRR-S4C4	25718	5.7753	-7471A	PPB		31 Jan 2017 18:29:12	U	SMPL
I1552-01 - 1	FK-EF-01-003-ABCD	426	0.0505	-7471A	PPB		31 Jan 2017 18:31:28	U	SMPL
I1552-01DUP - 1	FK-EF-01-003-ABCDUP	462	0.0586	-7471A	PPB		31 Jan 2017 18:36:22	U	SMPL
I1552-01MS - 1	FK-EF-01-003-ABCDMS	18501	4.1418	-7471A	PPB		31 Jan 2017 18:38:38	U	SMPL
I1552-01MSD - 1	FK-EF-01-003-ABCDMSD	18771	4.2029	-7471A	PPB		31 Jan 2017 18:40:55	U	SMPL
I1553-01 - 1	FK-BPM-01-A1B-A2B-A3B-A4B	232	0.0066	-7471A	PPB		31 Jan 2017 18:43:12	U	SMPL
I1554-01 - 1	FK-BPM-01-AA1-AA2-AA3-AA4	206	0.0007	-7471A	PPB		31 Jan 2017 18:45:30	U	SMPL
I1555-01 - 1	FK-BPM-01-AA5-AA6-AA7-AA8B	214	0.0025	-7471A	PPB		31 Jan 2017 18:47:48	U	SMPL
I1569-05 - 1	ELT-SB-03-10-12	281	0.0177	-7471A	PPB		31 Jan 2017 18:50:04	U	SMPL
CCV44 - 1	CCV44	22968	5.1529	-7471A	PPB		31 Jan 2017 18:52:20	U	SMPL
CCB44 - 1	CCB44	-31	-0.0530	-7471A	PPB		31 Jan 2017 18:54:36	U	SMPL
I1572-02 - 1	COMPOSITE-2	1128	0.2094	-7471A	PPB		31 Jan 2017 18:56:53	U	SMPL
I1593-03 - 1	B2-6	114346	25.8362	-7471A	PPB		31 Jan 2017 18:59:09	U	SMPL
I1593-06 - 1	B2-7	140402	31.7340	-7471A	PPB		31 Jan 2017 19:01:25	U	SMPL
I1593-09 - 1	B2-8	336546	HIGH	-7471A	PPB		31 Jan 2017 19:03:43	X	
I1552-01LX5 - 1		61	-0.0321	-7471A	PPB		31 Jan 2017 19:08:26	U	SMPL
I1552-01A - 1		18082	4.0469	-7471A	PPB		31 Jan 2017 19:10:43	U	SMPL
PB96556BS - 1	LCSS	17208	3.8491	-7471A	PPB		31 Jan 2017 19:14:43	U	SMPL
I1593-03DLX5 - 1		23332	5.2352	-7471A	PPB		31 Jan 2017 19:17:00	U	SMPL
I1593-06DLX5 - 1		28466	6.3973	-7471A	PPB		31 Jan 2017 19:19:18	U	SMPL
I1593-09DLX10 - 1		46338	10.4426	-7471A	PPB		31 Jan 2017 19:21:34	U	SMPL
CCV45 - 1	CCV45	21728	4.8722	-7471A	PPB		31 Jan 2017 19:23:51	U	SMPL
CCB45 - 1	CCB45	-31	-0.0530	-7471A	PPB		31 Jan 2017 19:26:07	U	SMPL
I1593-09DL2X50 - 1		9248	2.0473	-7471A	PPB		31 Jan 2017 19:31:01	U	SMPL
CCV46 - 1	CCV46	22854	5.1271	-7471A	PPB		31 Jan 2017 19:33:16	U	SMPL
CCB46 - 1	CCB46	-36	-0.0541	-7471A	PPB		31 Jan 2017 19:35:33	U	SMPL