


7470A

LB 85625
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36966



A= 0.0000e+000

B= 2.7045e-004

C= -2.1484e-002

Rho= 0.9999772

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	-3	0.000	-3				
0.2	0.200	0.181	-0.019	750	0.0 %	750				
2.5	2.500	2.531	0.031	9439	0.0 %	9439				
5.0	5.000	5.036	0.036	18701	0.0 %	18701				
7.5	7.500	7.497	-0.003	27801	0.0 %	27801				
10.0	10.000	9.976	-0.024	36966	0.0 %	36966				

LB85625 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Feb 2017 12:57:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-3	-	0.0000	7470A	PPB	01 Feb 2017 13:07:49	S	Std
0.2 - 1	50.2	750	-	0.2000	7470A	PPB	01 Feb 2017 13:09:57	S	Std
2.5 - 1	52.5	9439	-	2.5000	7470A	PPB	01 Feb 2017 13:12:06	S	Std
5.0 - 1	55	18701	-	5.0000	7470A	PPB	01 Feb 2017 13:14:14	S	Std
7.5 - 1	57.5	27801	-	7.5000	7470A	PPB	01 Feb 2017 13:16:22	S	Std
10.0 - 1	510	36966	-	10.0000	7470A	PPB	01 Feb 2017 13:18:29	S	Std
ICV14 - 1	ICV14	14671	3.9463	-7470A	PPB	01 Feb 2017 13:21:17	U	SMPL	
ICB14 - 1	ICB14	-48	-0.0345	-7470A	PPB	01 Feb 2017 13:23:23	U	SMPL	
CCV47 - 1	CCV47	19133	5.1531	-7470A	PPB	01 Feb 2017 13:25:29	U	SMPL	
CCB47 - 1	CCB47	-46	-0.0339	-7470A	PPB	01 Feb 2017 13:27:36	U	SMPL	
CRA - 1	CRA	745	0.1800	-7470A	PPB	01 Feb 2017 13:29:42	U	SMPL	
HighStd - 1	HighStd	36698	9.9036	-7470A	PPB	01 Feb 2017 13:31:49	U	SMPL	
ChkStd - 1	ChkStd	25177	6.7877	-7470A	PPB	01 Feb 2017 13:33:56	U	SMPL	
PB96602BL - 1	PBW	11	-0.0185	-7470A	PPB	01 Feb 2017 13:36:03	U	SMPL	
PB96602BS - 1	LCSW	14650	3.9406	-7470A	PPB	01 Feb 2017 13:38:11	U	SMPL	
PB96602TB - 1	PB96553TB	42	-0.0101	-7470A	PPB	01 Feb 2017 13:40:18	U	SMPL	
I1572-03 - 1	COMPOSITE-2	46	-0.0090	-7470A	PPB	01 Feb 2017 13:42:27	U	SMPL	
I1572-03DUP - 1	COMPOSITE-2DUP	69	-0.0028	-7470A	PPB	01 Feb 2017 13:44:35	U	SMPL	
I1572-03MS - 1	COMPOSITE-2MS	13385	3.5985	-7470A	PPB	01 Feb 2017 13:46:43	U	SMPL	
I1572-03MSD - 1	COMPOSITE-2MSD	13715	3.6878	-7470A	PPB	01 Feb 2017 13:48:50	U	SMPL	
CCV48 - 1	CCV48	19469	5.2439	-7470A	PPB	01 Feb 2017 13:50:57	U	SMPL	
CCB48 - 1	CCB48	-59	-0.0374	-7470A	PPB	01 Feb 2017 13:53:03	U	SMPL	
I1588-04 - 1	WC-B15-B21-001	96	0.0045	-7470A	PPB	01 Feb 2017 13:55:10	U	SMPL	
I1593-10 - 1	B2-6	113	0.0091	-7470A	PPB	01 Feb 2017 13:57:17	U	SMPL	
I1593-11 - 1	B2-7	104	0.0066	-7470A	PPB	01 Feb 2017 13:59:23	U	SMPL	
I1593-12 - 1	B2-8	112	0.0088	-7470A	PPB	01 Feb 2017 14:01:31	U	SMPL	
PB96603BL - 1	PBW	64	-0.0042	-7470A	PPB	01 Feb 2017 14:03:38	U	SMPL	
PB96603BS - 1	LCSW	14399	3.8728	-7470A	PPB	01 Feb 2017 14:05:46	U	SMPL	
I1529-06 - 1	ELT-FB12617	53	-0.0072	-7470A	PPB	01 Feb 2017 14:07:54	U	SMPL	
I1569-01 - 1	ELT-GW-01	443	0.0983	-7470A	PPB	01 Feb 2017 14:10:02	U	SMPL	
I1569-01DUP - 1	ELT-GW-01DUP	403	0.0875	-7470A	PPB	01 Feb 2017 14:12:11	U	SMPL	
I1569-02 - 1	I1569-01MS	16510	4.4437	-7470A	PPB	01 Feb 2017 14:14:18	U	SMPL	
CCV49 - 1	CCV49	18873	5.0828	-7470A	PPB	01 Feb 2017 14:16:25	U	SMPL	
CCB49 - 1	CCB49	-48	-0.0345	-7470A	PPB	01 Feb 2017 14:18:32	U	SMPL	
I1569-03 - 1	I1569-01MSD	16766	4.5129	-7470A	PPB	01 Feb 2017 14:20:38	U	SMPL	
I1569-04 - 1	ELT-GW-101	325	0.0664	-7470A	PPB	01 Feb 2017 14:22:45	U	SMPL	
I1594-01 - 1	MH-6	115	0.0096	-7470A	PPB	01 Feb 2017 14:24:52	U	SMPL	
I1572-03LX5 - 1		11	-0.0185	-7470A	PPB	01 Feb 2017 14:26:59	U	SMPL	
I1572-03A - 1		15553	4.1849	-7470A	PPB	01 Feb 2017 14:29:07	U	SMPL	
I1569-01LX5 - 1		64	-0.0042	-7470A	PPB	01 Feb 2017 14:31:15	U	SMPL	
I1569-01A - 1		16383	4.4093	-7470A	PPB	01 Feb 2017 14:33:23	U	SMPL	
CCV50 - 1	CCV50	19253	5.1855	-7470A	PPB	01 Feb 2017 14:38:50	U	SMPL	
CCB50 - 1	CCB50	-35	-0.0310	-7470A	PPB	01 Feb 2017 14:40:58	U	SMPL	