

ISM02.3

LB85632

~~LB85583~~

INSTRUMENT ID: CV1


NB

2/03/17

Linear

 μ Abs.:

43777



A= 0.0000e+000

B= 2.3098e-004

C= 1.4453e-003

Rho= 0.9997904

Accept=Accepted

slope
+/- intercept

td ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b D
00	0.000	0.003	0.003	5	0.000	5					
05	0.050	0.057	0.007	239	0.0 %	239					14
2	0.200	0.212	0.012	912	0.0 %	912					6
5	2.500	2.541	0.041	10994	0.0 %	10994					2
0	5.000	4.985	-0.015	21574	0.0 %	21574					0
5	7.500	7.340	-0.160	31772	0.0 %	31772					-2
1.0	10.000	10.113	0.113	43777	0.0 %	43777					1

NB
2/18/17NB
1/30/17

MS
2/8/17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		5	-		0 ISM02.3	1	1/30/2017 15:24	Std
0.05 S0.05		239	-	0.05	ISM02.3	1	1/30/2017 15:26	Std
0.2 S0.2		912	-	0.2	ISM02.3	1	1/30/2017 15:28	Std
2.5 S2.5		10994	-	2.5	ISM02.3	1	1/30/2017 15:30	Std
5 S5		21574	-	5	ISM02.3	1	1/30/2017 15:32	Std
7.5 S7.5		31772	-	7.5	ISM02.3	1	1/30/2017 15:34	Std
10 S10		43777	-	10	ISM02.3	1	1/30/2017 15:39	Std
ICV	ICV	16908	3.9069 -		ISM02.3	1	1/30/2017 15:43	SMPL
ICB	ICB	-22	-0.0036 -		ISM02.3	1	1/30/2017 15:45	SMPL
CCV033	CCV033	21301	4.9216 -		ISM02.3	1	1/30/2017 15:47	SMPL
CCB033	CCB033	-8	-0.0004 -		ISM02.3	1	1/30/2017 15:49	SMPL
PB96359BL	PBW003	3	0.0021 -		ISM02.3	1	1/30/2017 15:54	SMPL
I1390-11	MBDN53	90	0.0222 -		ISM02.3	1	1/30/2017 15:56	SMPL
I1390-12	MBDN54	74	0.0185 -		ISM02.3	1	1/30/2017 15:58	SMPL
I1390-13	MBDN55	76	0.019 -		ISM02.3	1	1/30/2017 16:00	SMPL
I1390-14	MBDN56	85	0.0211 -		ISM02.3	1	1/30/2017 16:02	SMPL
I1454-01	MYA913	732	0.1705 -		ISM02.3	1	1/30/2017 16:04	SMPL
I1454-02	MYA915	765	0.1781 -		ISM02.3	1	1/30/2017 16:07	SMPL
I1454-03	MYA917	448	0.1049 -		ISM02.3	1	1/30/2017 16:09	SMPL
I1454-04	MYA923	210	0.05 -		ISM02.3	1	1/30/2017 16:11	SMPL
I1454-05	MYA925	990	0.2301 -		ISM02.3	1	1/30/2017 16:13	SMPL
I1454-06	MYA925D	966	0.2246 -		ISM02.3	1	1/30/2017 16:15	SMPL
I1454-07	MYA925S	5131	1.1866 -		ISM02.3	1	1/30/2017 16:17	SMPL
I1454-08	MYA927	5310	1.2279 -		ISM02.3	1	1/30/2017 16:19	SMPL
I1454-09	MYA929	19048	4.4012 -		ISM02.3	1	1/30/2017 16:21	SMPL
I1454-10	MYA951	281	0.0664 -		ISM02.3	1	1/30/2017 16:24	SMPL
I1454-11	MYA953	483	0.113 -		ISM02.3	1	1/30/2017 16:26	SMPL
I1454-12	MYA967	80	0.0199 -		ISM02.3	1	1/30/2017 16:28	SMPL
I1454-13	MYA969	930	0.2163 -		ISM02.3	1	1/30/2017 16:30	SMPL
PB96361BL	PBW003	-4	0.0005 -		ISM02.3	1	1/30/2017 16:32	SMPL
I1455-01	MYA914	244	0.0578 -		ISM02.3	1	1/30/2017 16:34	SMPL
CCV034	CCV034	21213	4.9012 -		ISM02.3	1	1/30/2017 16:36	SMPL
CCB034	CCB034	-15	-0.002 -		ISM02.3	1	1/30/2017 16:38	SMPL
I1455-02	MYA916	159	0.0382 -		ISM02.3	1	1/30/2017 16:41	SMPL
I1455-03	MYA918	128	0.031 -		ISM02.3	1	1/30/2017 16:43	SMPL
I1455-04	MYA924	116	0.0282 -		ISM02.3	1	1/30/2017 16:45	SMPL
I1455-05	MYA926	249	0.059 -		ISM02.3	1	1/30/2017 16:47	SMPL
I1455-06	MYA926D	254	0.0601 -		ISM02.3	1	1/30/2017 16:49	SMPL
I1455-07	MYA926S	4018	0.9295 -		ISM02.3	1	1/30/2017 16:51	SMPL
I1455-08	MYA928	1476	0.3424 -		ISM02.3	1	1/30/2017 16:53	SMPL
I1455-09	MYA930	398	0.0934 -		ISM02.3	1	1/30/2017 16:55	SMPL
I1455-10	MYA952	167	0.04 -		ISM02.3	1	1/30/2017 16:57	SMPL
I1455-11	MYA954	255	0.0603 -		ISM02.3	1	1/30/2017 17:00	SMPL
I1455-12	MYA968	1	0.0017 -		ISM02.3	1	1/30/2017 17:02	SMPL
I1455-13	MYA970	528	0.1234 -		ISM02.3	1	1/30/2017 17:04	SMPL
I1462-05	MBDPB3	-22	-0.0036 -		ISM02.3	1	1/30/2017 17:06	SMPL

MS
2/08/17

LB85583 2385632
INSTRUMENT ID : CV1

MB
2/8/17

I1463-19	MBDMZ9	-20	-0.0032 -	ISM02.3	1	1/30/2017 17:08	SMPL
I1463-20	MBDN00	-15	-0.002 -	ISM02.3	1	1/30/2017 17:10	SMPL
I1501-16	MBDMX3	62	0.0158 -	ISM02.3	1	1/30/2017 17:12	SMPL
I1501-17	MBDMX4	32	0.0088 -	ISM02.3	1	1/30/2017 17:14	SMPL
I1501-18	MBDN57	17	0.0054 -	ISM02.3	1	1/30/2017 17:17	SMPL
CCV035	CCV035	21042	4.8617 -	ISM02.3	1	1/30/2017 17:19	SMPL
CCB035	CCB035	-15	-0.002 -	ISM02.3	1	1/30/2017 17:21	SMPL