



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85647

Temp.: 24.0 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/02/2017	12:10
2	CAL2	1	Water			7.00	02/02/2017	12:11
3	CAL3	1	Water			10.04	02/02/2017	12:12
4	ICV	1	Water			6.99	02/02/2017	12:13
5	CCV1	1	Water			2.01	02/02/2017	12:14
6	I1660-01	1	Solid	20.02	20	5.16	02/02/2017	12:15
7	I1660-01DUP	1	Solid	20.04	20	5.17	02/02/2017	12:16
8	I1660-02	1	Solid	20.02	20	7.25	02/02/2017	12:17
9	I1660-03	1	Solid	20.00	20	6.73	02/02/2017	12:18
10	I1660-04	1	Solid	20.02	20	7.71	02/02/2017	12:20
11	I1660-05	1	Solid	20.05	20	7.10	02/02/2017	12:21
12	I1660-06	1	Solid	20.02	20	6.64	02/02/2017	12:22
13	I1660-07	1	Solid	20.01	20	7.49	02/02/2017	12:23
14	I1660-08	1	Solid	20.03	20	8.16	02/02/2017	12:24
15	I1660-09	1	Solid	20.06	20	6.83	02/02/2017	12:25
16	CCV2	1	Water			12.02	02/02/2017	12:26
17	I1660-10	1	Solid	20.04	20	6.90	02/02/2017	12:27
18	I1660-11	1	Solid	20.03	20	7.90	02/02/2017	12:28
19	I1660-12	1	Solid	20.01	20	6.82	02/02/2017	12:30
20	CCV3	1	Water			12.01	02/02/2017	12:31

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY i1660

WorkList ID : 95272

Date : 2/2/2017 11:42:54 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1660-01	Corrosivity	Cool 4 deg C	LIRO01	G51	SB-34-COMP	01/30/2017	9045C
	Solid	I1660-02	Corrosivity	Cool 4 deg C	LIRO01	G51	SB-28-COMP	01/30/2017	9045C
	Solid	I1660-03	Corrosivity	Cool 4 deg C	LIRO01	G51	SB-22-COMP	01/30/2017	9045C
	Solid	I1660-04	Corrosivity	Cool 4 deg C	LIRO01	G51	SB-43-COMP	01/30/2017	9045C
	Solid	I1660-05	Corrosivity	Cool 4 deg C	LIRO01	G51	SB-44-COMP	01/30/2017	9045C
	Solid	I1660-06	Corrosivity	Cool 4 deg C	LIRO01	G51	SB-35-COMP	01/31/2017	9045C
	Solid	I1660-07	Corrosivity	Cool 4 deg C	LIRO01	G51	SB-36-COMP	01/31/2017	9045C
	Solid	I1660-08	Corrosivity	Cool 4 deg C	LIRO01	G51	SB-37-COMP	01/31/2017	9045C
	Solid	I1660-09	Corrosivity	Cool 4 deg C	LIRO01	G51	SB-38-COMP	01/31/2017	9045C
	Solid	I1660-10	Corrosivity	Cool 4 deg C	LIRO01	G51	SB-39-COMP	01/31/2017	9045C
	Solid	I1660-11	Corrosivity	Cool 4 deg C	LIRO01	G51	SB-41-COMP	01/31/2017	9045C
	Solid	I1660-12	Corrosivity	Cool 4 deg C	LIRO01	G51	SB-42-COMP	01/31/2017	9045C

Date/Time 02-02-17 11:42 AM

Received by:

Relinquished by:

Date/Time 02-02-17 6:00 PM

Received by:

Relinquished by: