

7470A


LB 8565D

INSTRUMENT ID: CV4

Linear

 μ Abs.:

39581



A= 0.0000e+000

B= 2.5553e-004

C= -2.3382e-002

Rho= 0.9996802

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	82	0.000	82				
0.2	0.200	0.200	0.000	876	0.0 %	876				
2.5	2.500	2.516	0.016	9937	0.0 %	9937				
5.0	5.000	5.086	0.086	19995	0.0 %	19995				
7.5	7.500	7.309	-0.191	28696	0.0 %	28696				
10.0	10.000	10.091	0.091	39581	0.0 %	39581				

LB85650 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Feb 2017 15:25:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	82	-	0.0000	7470A	PPB	02 Feb 2017 15:29:26	S	Std
0.2 - 1	30.2	876	-	0.2000	7470A	PPB	02 Feb 2017 15:31:35	S	Std
2.5 - 1	52.5	9937	-	2.5000	7470A	PPB	02 Feb 2017 15:33:43	S	Std
5.0 - 1	55	19995	-	5.0000	7470A	PPB	02 Feb 2017 15:35:51	S	Std
7.5 - 1	57.5	28696	-	7.5000	7470A	PPB	02 Feb 2017 15:37:59	S	Std
10.0 - 1	510	39581	-	10.0000	7470A	PPB	02 Feb 2017 15:40:06	S	Std
ICV15 - 1	ICV15	15208	3.8627	-	7470A	PPB	02 Feb 2017 15:42:39	U	SMPL
ICB15 - 1	ICB15	27	-0.0165	-	7470A	PPB	02 Feb 2017 15:44:45	U	SMPL
CCV51 - 1	CCV51	19194	4.8813	-	7470A	PPB	02 Feb 2017 15:46:52	U	SMPL
CCB51 - 1	CCB51	6	-0.0218	-	7470A	PPB	02 Feb 2017 15:48:58	U	SMPL
CRA - 1	CRA	873	0.1997	-	7470A	PPB	02 Feb 2017 15:51:05	U	SMPL
HighStd - 1	HighStd	38583	9.8358	-	7470A	PPB	02 Feb 2017 15:53:11	U	SMPL
ChkStd - 1	ChkStd	25669	6.5359	-	7470A	PPB	02 Feb 2017 15:55:17	U	SMPL
PB96623BL - 1	PBW	28	-0.0162	-	7470A	PPB	02 Feb 2017 15:57:25	U	SMPL
PB96623BS - 1	LCSW	13640	3.4621	-	7470A	PPB	02 Feb 2017 15:59:33	U	SMPL
PB96623TB - 1	PB96570TB	65	-0.0068	-	7470A	PPB	02 Feb 2017 16:01:40	U	SMPL
I1597-21 - 1	SB-26-COMP	85	-0.0017	-	7470A	PPB	02 Feb 2017 16:03:49	U	SMPL
I1597-21DUP - 1	SB-26-COMPDUP	103	0.0029	-	7470A	PPB	02 Feb 2017 16:05:57	U	SMPL
I1597-21MS - 1	SB-26-COMPMS	12933	3.2814	-	7470A	PPB	02 Feb 2017 16:08:06	U	SMPL
I1597-21MSD - 1	SB-26-COMPMSD	12729	3.2293	-	7470A	PPB	02 Feb 2017 16:10:12	U	SMPL
CCV52 - 1	CCV52	19281	4.9035	-	7470A	PPB	02 Feb 2017 16:12:19	U	SMPL
CCB52 - 1	CCB52	11	-0.0206	-	7470A	PPB	02 Feb 2017 16:14:25	U	SMPL
I1597-22 - 1	SB-27-COMP	135	0.0111	-	7470A	PPB	02 Feb 2017 16:16:32	U	SMPL
I1597-23 - 1	SB-30-COMP	155	0.0162	-	7470A	PPB	02 Feb 2017 16:18:39	U	SMPL
I1597-24 - 1	SB-25-COMP	156	0.0165	-	7470A	PPB	02 Feb 2017 16:20:46	U	SMPL
I1597-25 - 1	SB-29-COMP	145	0.0137	-	7470A	PPB	02 Feb 2017 16:22:53	U	SMPL
I1597-26 - 1	SB-31-COMP	160	0.0175	-	7470A	PPB	02 Feb 2017 16:25:00	U	SMPL
I1597-27 - 1	SB-32-COMP	88	-0.0009	-	7470A	PPB	02 Feb 2017 16:27:08	U	SMPL
I1597-28 - 1	SB-33-COMP	77	-0.0037	-	7470A	PPB	02 Feb 2017 16:29:16	U	SMPL
I1597-29 - 1	SB-50-COMP	58	-0.0086	-	7470A	PPB	02 Feb 2017 16:31:25	U	SMPL
I1597-30 - 1	SB-40-COMP	66	-0.0065	-	7470A	PPB	02 Feb 2017 16:33:33	U	SMPL
I1644-05 - 1	WC-B1-B2-001	268	0.0451	-	7470A	PPB	02 Feb 2017 16:35:40	U	SMPL
CCV53 - 1	CCV53	19556	4.9738	-	7470A	PPB	02 Feb 2017 16:37:47	U	SMPL
CCB53 - 1	CCB53	28	-0.0162	-	7470A	PPB	02 Feb 2017 16:39:54	U	SMPL
I1644-06 - 1	WC-B3-B5-001	199	0.0275	-	7470A	PPB	02 Feb 2017 16:42:01	U	SMPL
I1597-21LX5 - 1		75	-0.0042	-	7470A	PPB	02 Feb 2017 16:44:08	U	SMPL
I1597-21A - 1		15174	3.8541	-	7470A	PPB	02 Feb 2017 16:46:15	U	SMPL
CCV54 - 1	CCV54	19092	4.8552	-	7470A	PPB	02 Feb 2017 16:48:22	U	SMPL
CCB54 - 1	CCB54	28	-0.0162	-	7470A	PPB	02 Feb 2017 16:50:29	U	SMPL