


245.1

LB 85654
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36550



A= 0.0000e+000

B= 2.7499e-004

C= -3.9788e-002

Rho= 0.9998956

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.029	-0.029	41	0.000	41				
0.2	0.200	0.162	-0.038	732	0.0 %	732				
2.5	2.500	2.610	0.110	9636	0.0 %	9636				
5.0	5.000	4.996	-0.004	18312	0.0 %	18312				
7.5	7.500	7.450	-0.050	27236	0.0 %	27236				
10.0	10.000	10.011	0.011	36550	0.0 %	36550				

LB85654 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 02 Feb 2017 17:22:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	41	-	0.0000	245.1	PPB	02 Feb 2017 17:23:42	S	Std
0.2 - 1	50.2	732	-	0.2000	245.1	PPB	02 Feb 2017 17:25:50	S	Std
2.5 - 1	82.5	9636	-	2.5000	245.1	PPB	02 Feb 2017 17:27:59	S	Std
5.0 - 1	85	18312	-	5.0000	245.1	PPB	02 Feb 2017 17:30:07	S	Std
7.5 - 1	87.5	27236	-	7.5000	245.1	PPB	02 Feb 2017 17:32:14	S	Std
10.0 - 1	87.5	36550	-	10.0000	245.1	PPB	02 Feb 2017 17:34:21	S	Std
ICV16 - 1	ICV16	15130	4.1208	-245.1		PPB	02 Feb 2017 17:36:59	U	SMPL
ICB16 - 1	ICB16	13	-0.0362	-245.1		PPB	02 Feb 2017 17:39:05	U	SMPL
CCV55 - 1	CCV55	17888	4.8793	-245.1		PPB	02 Feb 2017 17:41:12	U	SMPL
CCB55 - 1	CCB55	7	-0.0379	-245.1		PPB	02 Feb 2017 17:43:18	U	SMPL
CRA - 1	CRA	793	0.1783	-245.1		PPB	02 Feb 2017 17:45:24	U	SMPL
HighStd - 1	HighStd	36268	9.9336	-245.1		PPB	02 Feb 2017 17:47:31	U	SMPL
ChkStd - 1	ChkStd	25068	6.8537	-245.1		PPB	02 Feb 2017 17:49:38	U	SMPL
PB96622BL - 1	PBW	109	-0.0098	-245.1		PPB	02 Feb 2017 17:51:45	U	SMPL
PB96622BS - 1	LCSW	14438	3.9306	-245.1		PPB	02 Feb 2017 17:53:53	U	SMPL
I1641-01 - 1	013017-DSDC-F-D-M	467	0.0886	-245.1		PPB	02 Feb 2017 17:56:00	U	SMPL
I1641-02 - 1	013017-DSDC-F-U-M	562	0.1148	-245.1		PPB	02 Feb 2017 17:58:09	U	SMPL
I1641-02DUP - 1	013017-DSDC-F-U-MDUP	647	0.1381	-245.1		PPB	02 Feb 2017 18:00:17	U	SMPL
I1641-02MS - 1	013017-DSDC-F-U-MMS	16306	4.4442	-245.1		PPB	02 Feb 2017 18:02:26	U	SMPL
I1641-02MSD - 1	013017-DSDC-F-U-MMSD	15582	4.2451	-245.1		PPB	02 Feb 2017 18:04:33	U	SMPL
CCV56 - 1	CCV56	18144	4.9497	-245.1		PPB	02 Feb 2017 18:06:39	U	SMPL
CCB56 - 1	CCB56	9	-0.0373	-245.1		PPB	02 Feb 2017 18:08:46	U	SMPL
I1656-01 - 1	COMP	302	0.0433	-245.1		PPB	02 Feb 2017 18:10:53	U	SMPL
I1641-02LX5 - 1		137	-0.0021	-245.1		PPB	02 Feb 2017 18:13:00	U	SMPL
I1641-02A - 1		15131	4.1211	-245.1		PPB	02 Feb 2017 18:15:07	U	SMPL
CCV57 - 1	CCV57	18301	4.9928	-245.1		PPB	02 Feb 2017 18:17:14	U	SMPL
CCB57 - 1	CCB57	14	-0.0359	-245.1		PPB	02 Feb 2017 18:19:21	U	SMPL